Bilirubin Supplemental, 2 vial

Name	Line No.	Specimen 1 Range & Type	Mean	SD	Specimen 2 Range & Type	Mean	SD	Specimen 3 Range & Type	Mean	SD	Specimen 4 Range & Type	Mean	SD	Specimen 5 Range & Type	Mean	SD	No. of Labs
Bilirubin, Direct																	
Initial Grouping by Reagent and Instrument																	
Beckman Olympus & Beck Olym AU 400/600/5400	1	4.1 - 6.1	P 5.11	0.51	0 - 0.4	C 0.0	0.0	-			-			-			11
Ortho Vitros & Ortho Vitros not DT or ECi	2	3.2 - 4.9	P 4.06	0.23	0 - 0.4	C 0.0	0.0	-			-			-			11
Siemens Dimension DBI & Siemens Dimension EXL	3	2.8 - 4.2	P 3.53	0.24	0 - 0.5	C 0.07	0.05	-			-			-			36
Initial Grouping by Reagent																	
Beckman Olympus	4	4.2 - 6.2	P 5.19	0.42	0 - 0.4	C 0.0	0.0	-			-			-			20
Ortho Vitros	5	3.3 - 4.9	P 4.07	0.26	0 - 0.4	C 0.0	0.0	-			-			-			19
Roche acid diazo	6	3.1 - 4.7	P 3.92	0.27	0 - 0.5	C 0.09	0.1	-			-			-			16
Siemens Dimension DBI	7	2.8 - 4.2	P 3.52	0.23	0 - 0.5	C 0.07	0.05	-			-			-			46
Initial Grouping by Sensitivity or Principle																	
Acid diazo methods	8	3.1 - 4.6	P 3.82	0.67	0 - 0.5	C 0.07	0.07	-			-			-			112
Diazonium ion methods	9	4.1 - 6.1	P 5.1	0.52	0 - 0.4	C 0.0	0.02	-			-			-			23
Total Population																	
Whole Population	10	3.2 - 4.8	P 4.04	0.8	0 - 0.5	C 0.06	0.07	-			-			-			137
Bilirubin, Neonatal																	
Initial Grouping by Reagent																	
Ortho Vitros	1	12.6 - 18.8	P 15.7	0.56	0 - 0.4	C 0.01	0.03	-			-			-			10
Initial Grouping by Sensitivity or Principle																	
Diazonium ion	2	12.6 - 18.8	P 15.7	0.56	0 - 0.4	C 0.01	0.03	-			-			-			10
Total Population																	
Whole Population	3	12.6 - 18.8	P 15.69	0.5	0 - 0.4	C 0.04	0.05	-			-			-			22
Bilirubin, Total supplemenal																	
Initial Grouping by Reagent and Instrument																	
Siemens Dimension & Siemens Dimension EXL	1	12.5 - 18.7	P 15.59	0.47	0 - 0.5	C 0.09	0.05	-			-			-			17
Initial Grouping by Reagent																	
Ortho Vitros	2	12.9 - 19.3	P 16.11	0.61	0 - 0.5	C 0.09	0.05	-			-			-			11
Siemens Dimension	3	12.5 - 18.8	P 15.63	0.61	0 - 0.5	C 0.08	0.05	-			-			-			23
Initial Grouping by Sensitivity or Principle																	
Diazonium ion	4	12.5 - 18.8	P 15.64	0.84	0 - 0.5	C 0.09	0.06	-			-			-			23
Diazo/caffeine-benzoate	5	12.5 - 18.8	P 15.67	0.57	0 - 0.5	C 0.08	0.05	-			-			-			28
Total Population																	
Whole Population	6	12.5 - 18.8	P 15.65	0.8	0 - 0.5	C 0.08	0.06	-			-			-			62