D-Dimer

		Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			
Name	**Line No.**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**No. of Labs**
D-Dimer, quant ng/mL																	
Initial Grouping by Reagent																	
Alere Triage	1	1494 - 2828	S 2160.9	222.4	460 - 1006	S 732.9	90.9	-			-			-			57
Siemens Stratus CS	2	2677 - 3376	S 3026.5	116.5	797 - 1071	S 934.1	45.7	-			-			-			10
Initial Grouping by Sensitivity or Principle																	
low recovery	3	705 - 3353	S 2028.9	441.4	288 - 1164	S 725.9	145.9	-			-			-			63
Standardized method	4	2677 - 3376	S 3026.5	116.5	0 - 1982	S 921.4	353.5	-			-			-			12
Total Population																	
Whole Population	5	480 - 3943	S 2211.7	577.1	117 - 1436	S 776.1	219.8	-			-			-			84
D-Dimer, quant ugFEU/mL																	
Initial Grouping by Reagent																	
Siemens Innovance	1	3.39 - 5.29	S 4.337	0.317	1.38 - 1.98	C 1.676	0.081	-			-			-			12
Initial Grouping by Sensitivity or Principle																	
mid recovery	2	1.55 - 6.55	S 4.052	0.833	0.47 - 2.66	S 1.564	0.365	-			-			-			19
Total Population																	
Whole Population	3	0.91 - 6.65	S 3.778	0.956	0.07 - 2.79	S 1.428	0.454	-			-			-			26