



**AMERICAN ASSOCIATION OF BIOANALYSTS
PROFICIENCY TESTING
11931 Wickchester Ln., Ste 200 - Houston, TX 77043
800-234-5315 - 281-436-5357 - Fax 713-781-5008**

SECOND EVENT 2022

Gastric Occult Blood

Occult Blood

Gastric Occult Blood	Specimen 6		Specimen 7	
	Neg	Pos	Neg	Pos
Beckman Coulter Gastroccult		5	5	
Total Population	0	5	5	0
Flagging	***			***

pH - Gastric	Specimen 6					Specimen 7				
	2.0	3.0	4.0	5.0	>5.0	2.0	3.0	4.0	5.0	>5.0
Beckman Coulter Gastroccult	3	2						1		2
Total Population	3	2	0	0	0	0	0	1	0	2
Flagging					***	***	***			

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.