



AMERICAN ASSOCIATION OF BIOANALYSTS
PROFICIENCY TESTING
 11931 Wickchester Ln., Ste 200 - Houston, TX 77043
 800-234-5315 - 281-436-5357 - Fax 713-781-5008

SECOND EVENT 2022

URINE MICROALBUMIN - CREATININE, qualitative

Microalbumin, semiquantative or qualitative

Name	Specimen 6									
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L
Siemens Clinitek Status	3		14						1	
Total Population	3	0	15	1	0	0	0	2	1	0
Flagging	***				***	***	***			

Name	Specimen 7									
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L
Siemens Clinitek Status									7	11
Total Population	0	0	0	2	1	0	0	0	7	12
Flagging	***	***				***	***	***		

Creatinine, semiquantitative

Name	Specimen 6					Specimen 7				
	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL
Siemens Clinitek Status			1	1	17				2	17
Total Population	0	0	1	2	18	0	0	0	3	18
Flagging	***	***				***	***			

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.