



## Ammonia, Serum

Name	Line No.	Specimen 1				Specimen 2				No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD			
<b>Ammonia, Serum</b>										
<b>Initial Grouping by Reagent and Instrument</b>										
Beckman Synchron & Beckman Unicel DXC series	1	65.0 - 85.0	C	74.9	4.8	100.0 - 122.0	P	110.8	4.1	21
J&J Vitros & J&J Vitros 5,1 FS	2	52.0 - 72.0	C	61.5	4.2	87.0 - 107.0	C	96.9	5.7	11
Siemens Dimension & Siemens Dimension Rxl	3	66.0 - 86.0	C	75.5	5.0	103.0 - 126.0	P	114.8	4.3	10
Siemens Dimension & Siemens Dimension EXL	4	64.0 - 84.0	C	73.6	6.6	102.0 - 125.0	P	113.7	3.8	12
Siemens Dimension & Siemens Dimension Xpand	5	66.0 - 86.0	C	76.5	6.7	103.0 - 126.0	P	114.3	4.7	18
<b>Initial Grouping by Reagent</b>										
Beckman Synchron	6	65.0 - 85.0	C	74.7	4.6	100.0 - 122.0	P	111.1	3.9	24
J&J Vitros	7	51.0 - 71.0	C	61.1	5.3	88.0 - 108.0	C	97.6	5.8	25
Siemens Dimension	8	65.0 - 85.0	C	75.0	5.9	103.0 - 126.0	P	114.1	4.2	49
<b>Initial Grouping by Sensitivity or Principle</b>										
Dry Slide or Cartridge	9	51.0 - 71.0	C	61.1	5.3	88.0 - 108.0	C	97.6	5.8	25
All other methods	10	65.0 - 85.0	C	74.9	7.4	101.0 - 123.0	P	112.2	9.8	92
<b>Total Population</b>										
Whole Population	11	62.0 - 82.0	C	71.7	7.6	98.0 - 119.0	P	108.4	8.1	124