



## Brain Natriuretic Peptide (BNP)

Name	Line No.	Specimen 1				Specimen 2				No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD			
<b>Brain Natriuretic Peptide (BNP)</b>										
<b>Initial Grouping by Reagent and Instrument</b>										
Abbott Architect & Abbott Architect c, ci, i	1	74.0 - 147.0	S	110.6	12.1	784.0 - 1184.0	S	983.9	66.7	16
Abbott AxSYM & Abbott AxSYM	2	122.0 - 319.0	S	220.1	32.8	647.0 - 1482.0	S	1064.1	139.2	15
Biosite Triage Card Panel & Biosite Triage Meter	3	354.0 - 477.0	S	415.5	20.6	1401.0 - 1748.0	S	1574.3	57.8	104
Roche ELECYSYS proBNP & Roche Elecsys series	4	276.0 - 517.0	S	396.2	40.1	1399.0 - 3041.0	S	2220.0	273.7	10
<b>Initial Grouping by Reagent</b>										
Abbott Architect	5	75.0 - 148.0	S	111.3	12.3	780.0 - 1178.0	S	978.9	66.3	16
Abbott AxSYM	6	124.0 - 309.0	S	216.4	30.9	638.0 - 1469.0	S	1053.3	138.6	15
Biosite Triage Card Panel	7	354.0 - 476.0	S	415.3	20.3	1401.0 - 1748.0	S	1574.3	57.8	107
Ortho Vitros NT-proBNP	8	814.0 - 1017.0	S	915.3	33.8	4643.0 - 5477.0	S	5060.0	139.1	12
Roche ELECYSYS proBNP	9	301.0 - 492.0	S	396.2	31.8	1579.0 - 2882.0	S	2230.4	217.0	18
Siemens	10	103.0 - 174.0	S	138.2	11.8	524.0 - 892.0	S	708.0	61.4	11
Siemens Dimension proBNP	11	0.0 - 721.0	S	238.7	160.8	0.0 - 3508.0	S	1549.0	652.8	12
Siemens NT-PBNP	12	0.0 - 1026.0	S	383.4	214.1	0.0 - 4848.0	S	2112.4	912.0	19
<b>Initial Grouping by Sensitivity or Principle</b>										
All BNP	13	0.0 - 2339.0	S	509.6	609.8	0.0 - 13691.0	S	2895.0	3598.8	15
All proBNP	14	0.0 - 1420.0	S	462.8	319.2	0.0 - 7675.0	S	2601.9	1691.1	90
All Biosite BNP	15	0.0 - 1122.0	S	482.7	212.9	0.0 - 5229.0	S	2014.9	1071.3	121
All Abbott BNP	16	0.0 - 347.0	S	173.6	57.7	691.0 - 1373.0	S	1032.1	113.6	38
<b>Total Population</b>										
Whole Population	17	389.0 - 459.0	S	424.1	11.7	2033.0 - 2378.0	S	2205.5	57.6	264