



D-Dimer

Name	Line No.	Specimen 1			Specimen 2			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	
D-Dimer, quantitative								
Initial Grouping by Method								
Biosite Triage	1	288.0 - 687.0 S	487.3	66.6	976.0 - 2190.0 S	1582.9	202.4	122
Diagnostica Stago Liatest	2	503.0 - 830.0 S	666.3	54.5	945.0 - 3076.0 S	2010.8	355.2	12
Siemens Stratus CS	3	575.0 - 916.0 S	745.8	56.8	2384.0 - 2997.0 S	2690.2	102.2	25
Siemens Sysmex Advanced	4	1123.0 - 1768.0 S	1445.3	107.4	3206.0 - 4959.0 S	4082.7	292.1	16
Initial Grouping by Sensitivity or Principle								
low recovery	5	288.0 - 687.0 S	487.3	66.6	976.0 - 2190.0 S	1582.9	202.4	122
mid recovery	6	1123.0 - 1768.0 S	1445.3	107.4	3206.0 - 4959.0 S	4082.7	292.1	16
high recovery	7	112.0 - 1296.0 S	704.2	197.4	753.0 - 4042.0 S	2397.4	548.2	45
IL Non ACL 7000	8	245.0 - 660.0 S	452.3	69.2	817.0 - 1203.0 S	1009.5	64.3	12
Total Population								
Whole Population	9	0.0 - 1459.0 S	607.9	283.8	0.0 - 4449.0 S	1926.2	841.0	214