



High Sensitivity C-Reactive Protein

Name	Line No.	Specimen 1			Specimen 2			No. of Labs		
		Range & Type	Mean	SD	Range & Type	Mean	SD			
High Sensitivity C-Reactive Protein										
Initial Grouping by Method										
Beckman Synchron turbidimet	1	2.06 - 2.89	S	2.475	0.139	9.39 - 10.28	S	9.834	0.148	11
Beckman Unicel	2	1.67 - 3.29	S	2.482	0.269	7.94 - 11.95	S	9.948	0.668	12
Carolina immunoturbid	3	2.02 - 2.94	S	2.481	0.154	8.89 - 12.15	S	10.521	0.543	15
Beckman Olympus turbid	4	0.14 - 4.18	S	2.162	0.673	0.0 - 19.9	S	8.237	3.889	11
Roche COBAS Integra	5	1.34 - 3.24	S	2.291	0.316	7.95 - 10.26	S	9.103	0.384	10
Siemens CCRP	6	2.29 - 4.26	S	3.272	0.328	6.98 - 14.56	S	10.768	1.263	19
Siemens Extended RCRP	7	2.11 - 3.36	S	2.732	0.208	9.9 - 11.31	S	10.603	0.236	11
Siemens Immulite 2000	8	1.66 - 3.32	S	2.489	0.278	7.97 - 13.16	S	10.565	0.864	10
Initial Grouping by Sensitivity or Principle										
All wide range methods	9	1.57 - 3.49	S	2.531	0.320	1.07 - 20.92	S	10.997	3.308	20
All Roche HS methods	10	1.59 - 3.07	S	2.331	0.247	7.64 - 10.8	S	9.218	0.527	23
All direct quantitative meth	11	1.17 - 3.68	S	2.425	0.419	4.76 - 15.15	S	9.956	1.731	106
All quantitations by dilutn	12	1.38 - 4.42	S	2.900	0.508	5.96 - 15.79	S	10.878	1.638	41
Total Population										
Whole Population	13	1.54 - 3.52	S	2.528	0.330	7.16 - 13.38	S	10.266	1.037	190