



**PROFICIENCY TESTING SERVICE  
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596  
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**THIRD QUADRIMESTER 2011**

## Acetone

Acetone	Specimen 1				Specimen 2				Specimen 3				Specimen 4				Specimen 5					
	Name	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	
Bayer Acetest	1	14	44	40	97	1	1			14	82	2	1	99				99				
Gas chromat				1	1						1			1				1				
<b>Total Population</b>	<b>1</b>	<b>15</b>	<b>47</b>	<b>43</b>	<b>104</b>	<b>1</b>	<b>1</b>			<b>15</b>	<b>87</b>	<b>3</b>	<b>1</b>	<b>106</b>				<b>106</b>				
<i>Flagging</i>	*****					*****	*****	*****	*****	*****					*****	*****	*****		*****	*****	*****	*****

Correct responses are defined as those reflecting agreement among 80% or more of all participants.  
Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.