



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. • Brownsville, TX 78520-5596
800-234-5315 • 281-436-5357 • Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2011

CARDIAC MARKERS / ISOENZYMES

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated
CK-2 (CK-MB)										
LifeSign MI	1		1			1	1		1	
Nanogen Cardiac STATus	4		4		3	1	4		4	
Total Population	6	0	6	0	4	2	6	0	6	0
Flagging		*****		*****				*****		*****

Due to a lack of participant consensus, Specimen 3 was not evaluated this event.

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated
Myoglobin, qualitative										
LifeSign MI		1	1			1	1		1	
Nanogen Cardiac STATus	2	2	4		4	1	3	1	4	
Total Population	3	3	6	0	0	6	5	1	6	0
Flagging				*****	*****			*****		*****

Due to a lack of participant consensus, Specimen 1 was not evaluated this event.

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated
Troponin I										
LifeSign MI	1		1			1	1		1	
Nanogen Cardiac STATus	4		4		1	3	4		4	
Total Population	5	0	5	0	1	4	5	0	5	0
Flagging		*****		*****	*****			*****		*****

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated
Troponin T										
Roche Trop T Sensitive		3	3			3	2	1	3	
Total Population	0	4	4	0	0	4	3	1	4	0
Flagging	*****			*****	*****					*****

Due to a lack of participant consensus, Specimen 4 was not evaluated this event.

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.