



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2011

D-Dimer, Qualitative

Name
American Diag. Simplify
bioMerieux Vidas D-Dimer
Biosite Triage
Diagnostica Stago Liatest
Inverness Clearview
Pacific Hemostasis
Remel Murex
Siemens Dimertest
Total Population
Flagging

Specimen 1	
Positive	Negative
1	
1	
5	1
2	
10	6
	1
	1
	1
20	10

Specimen 2	
Positive	Negative
1	
1	
5	1
2	
16	
1	
	1
1	
28	2

Due to a lack of participant consensus, Specimen 1 was not evaluated this event.

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.