



**PROFICIENCY TESTING SERVICE  
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596  
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**THIRD QUADRIMESTER 2011**

**Fecal Occult Blood**

Name
Beck/Coult Hemmocult
Beck/Coult Hemmocult ICT
Beck/Coult Hemmocult SENSAs
Cenogenics
Fischer Sure-View ES
Fischer Sure-View FOB
Helena ColoScreen
HemaPrompt
Hemosure Onestep
Henry Schein One-Step
Immunos Hema Screen Specific
Immunostics Hema Screen
Invern Clear iFOB Complete
Inverness Clearview Ultra
Polymedco OC-Light
PolymedcoOC-Auto Micro 80
Propper Seracult
Propper Seracult Plus
PSS Select
PSS Select-ES
Quidel IFOB
<b>Total Population</b>
<b>Flagging</b>

Specimen 1	
Positive	Negative
3	131
	3
1	128
	3
	11
	20
	31
	3
	33
	2
	2
	11
	7
	4
	5
	4
	34
	13
1	9
	2
	18
<b>5</b>	<b>480</b>
*****	

Specimen 2	
Positive	Negative
131	3
3	
129	
3	
11	
20	
31	
3	
29	4
2	
2	
11	
3	4
3	1
5	
4	
33	1
13	
9	1
2	
18	
<b>471</b>	<b>14</b>
	*****

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.