



**PROFICIENCY TESTING SERVICE  
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. • Brownsville, TX 78520-5596  
800-234-5315 • 281-436-5357 • Fax 713-781-5008

**PARTICIPANT STATISTICS**

**THIRD QUADRIMESTER 2011**

**Serum Pregnancy**

Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg
Alpha Sci Instant-View Cmbo	6					6	6					6	6		
Beckman ICON 20/25 s/u	27		1			28	28					28	27		
Beckman ICON II s/u	3					3	3					3	3		
Cardinal Health SP	4					4	4					4	4		
CardinalHealth hCG Combo	17					17	17					17	17		
Fisher Sure-Vue hCG-STAT	14	1				15	15					15	15		
Fisher Sure-Vue s/u hCG	17	1				18	18					18	18		
Genzyme OSOM hCG-Combo	9					9	9					9	9		
Henry Schein OneStep	9					10	9	1				10	10		
Inverness Acceava Combo	20				1	19	20		1	2	17	20	20		
McKesson Medi-Lab	12					12	12					12	12		
OneStep 3rd party, OEM	4					3	4				1	3	4		
OneStep 3rd party, OEM waiv							1					1	1		
Polymedco Poly stat hCG	8					8	8					8	8		
PSS Select Combo	3					3	3					3	3		
Quidel Quickvue OS Combo	18	1				19	19					19	19		
Quidel Quickvue+ OS Combo	22					22	21	1				22	22		
SA Scientific SAS	4					4	4					4	4		
Wampole Clearview Combo	4					4	4					4	4		
<b>Total Population</b>	<b>232</b>	<b>3</b>	<b>1</b>	<b>0</b>	<b>2</b>	<b>235</b>	<b>235</b>	<b>1</b>	<b>2</b>	<b>2</b>	<b>3</b>	<b>233</b>	<b>237</b>	<b>0</b>	<b>0</b>
<b>Flagging</b>		***	***	***	***			***	***		***	***		***	***

**Urine Pregnancy**

Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg
Alpha Sci Instant-View Cmbo	13					13	13					13	13		
Beckman ICON 20/25 s/u	47	1				48	47	1				48	47	1	
Beckman ICON II s/u	4					4	4					4	4		
Cardinal Health SP	14					14	14					14	14		
CardinalHealth hCG Combo	19					19	19					19	19		
Diag Sp Pregna-Cert	16	1				17	17			1	16	17	17		
Fisher Sure-Vue hCG-STAT	34	1				35	35					35	35		
Fisher Sure-Vue s/u hCG	19	2				21	20	1				21	20	1	
Fisher Sure-Vue urine hCG	6					6	6					6	6		
Genzyme OSOM hCG-Combo	22					22	22					22	22		
Genzyme OSOM hCG-Urine	17					17	17					17	17		
Henry Schein OneStep	11					11	11					11	11		
Inverness Acceava Combo	13					13	13					13	13		
Inverness BioStar Acceav II	4					4	4					4	4		
Inverness Clearview Ultra	6					6	6					6	6		
Lifesign Status s/u hCG	4					4	4					4	4		
McKesson Medi-Lab	16	2				18	18					18	18		
NCD Pro Advantage	6					6	6					6	6		
OneStep 3rd party, OEM waiv	4					4	4					4	4		
PerMaxim OneStep	4					4	4					4	4		
Polymedco Poly stat hCG	11					11	11					11	11		
PSS Consult	13					13	13					13	12		1
PSS Select Combo	8					8	8					8	8		
Qui QuickVue Urine/One-Step	68	1				69	68	1				69	68	1	
Quidel Quickvue OS Combo	33					33	33					33	33		
Quidel Quickvue+ OS Combo	35					35	34		1	1	33	35	35		
Stanbio QuPID	4					4	4					4	4		
Stanbio QuPID PLUS	5					5	5					5	5		
Stanbio True 20 Plus	12					12	12					12	12		
Wampole Clearview Combo	6					6	6					6	6		
<b>Total Population</b>	<b>533</b>	<b>8</b>	<b>1</b>	<b>0</b>	<b>0</b>	<b>540</b>	<b>536</b>	<b>4</b>	<b>1</b>	<b>1</b>	<b>1</b>	<b>538</b>	<b>537</b>	<b>3</b>	<b>1</b>
<b>Flagging</b>		***	***	***	***			***	***	***	***		***	***	***

*hcg Concentration*

3030 mIU/mL

<1 mIU/mL

3030 mIU/mL

<1 mIU/mL

3030 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.  
Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.