



**PROFICIENCY TESTING SERVICE  
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596  
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**URINE MICROALBUMIN - CREATININE, qualitative**

**THIRD QUADRIMESTER 2011**

**Microalbumin, semiquantative or qualitative**

Name	Specimen 1								Specimen 2							
	20 mg/L	30 mg/L	50 mg/L	80 mg/L	100 mg/L	150 mg/L	Negative	Positive	20 mg/L	30 mg/L	50 mg/L	80 mg/L	100 mg/L	150 mg/L	Negative	Positive
BTNX Rapid Response																
CLIAwaived Microalb 2-1	1										1					
Roche Chemstrip Micral	1		7	3	6			4	1		4		12			4
Siemens Clinitek 50		5		14									16		3	
Siemens Clinitek 100		4		4						1			7			
Siemens Clinitek Status		1		54		4		1					34	2	22	1
	2	10	7	76	6	4	0	5	1	1	5	58	14	25	0	5
<b>Flagging</b>	*****					*****	*****		*****	*****					*****	

**Creatinine, semiquantitative**

Name	Specimen 1					Specimen 2				
	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL
BTNX Rapid Response					1					1
Siemens Clinitek 100			1	7				2	6	
Siemens Clinitek 50			10	8	1			9	10	
Siemens Clinitek Status			17	34	3	1		15	31	6
Siemens Micro-Bumintest		1				1				
Total Population	0	1	28	50	5	0	2	26	48	7
<b>Flagging</b>	*****	*****				*****	*****			

Correct responses are defined as those reflecting agreement among 80% or more of all participants.  
Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.