



Glycohemoglobin, 5 vial

Name	Line No.	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	
Glycohemoglobin - HbA1c 5v																	
Initial Grouping by Reagent																	
Beckman Olympus	1	7.9 - 10.7	P 9.28	0.77	4.0 - 6.0	C 5.0	0.32	4.0 - 6.0	C 4.98	0.27	7.7 - 10.4	P 9.05	0.55	4.0 - 6.0	C 4.98	0.26	4
Initial Grouping by Sensitivity or Principle																	
Values standardized to DCCT	2	7.9 - 10.7	P 9.27	0.58	4.1 - 6.1	C 5.1	0.28	4.1 - 6.1	C 5.09	0.25	7.8 - 10.5	P 9.16	0.44	4.1 - 6.1	C 5.07	0.23	7
Liquid chromatography	3	8.2 - 11.2	P 9.7	0.45	4.6 - 6.6	C 5.57	0.26	4.7 - 6.7	C 5.73	0.05	8.4 - 11.4	P 9.93	0.29	4.8 - 6.8	C 5.77	0.05	4
Total Population																	
Whole Population	4	8.2 - 11.1	P 9.61	0.58	4.3 - 6.3	C 5.25	0.29	4.3 - 6.3	C 5.29	0.3	8.1 - 11.0	P 9.58	0.52	4.3 - 6.3	C 5.29	0.31	17
Glycohemoglobin - HbA1 5v																	
Total Population																	
Whole Population	1	0 - 2.0	C 1.0	0.0	0 - 1.4	C 0.4	0.0	0 - 1.4	C 0.4	0.0	0 - 2.0	C 1.0	0.0	0 - 1.4	C 0.4	0.0	1
Glycohemoglobin, Total 5v																	
Total Population																	
Whole Population	1	10.3 - 17.1	P 13.7	0.0	10.6 - 17.6	P 14.1	0.0	11.1 - 18.5	P 14.8	0.0	10.4 - 17.3	P 13.8	0.0	11.0 - 18.3	P 14.6	0.0	1