



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

THIRD QUADRIMESTER 2018

PREGNANCY, waived

Name	Specimen 1			Specimen 2		
	Neg	Pos	Ind	Neg	Pos	Ind
Alere		13		13		
CardinalHealth hCG Combo		21		21		
Consult Diagnostics		31		31		
Fisher Sure-Vue urine hCG		19		19		
Henry Schein OneStep		11		11		
Quidel QuickVue	1	17		18		
Quidel Quickvue+ OS Combo		23		23		
Sekisui OSOM hCG-Urine		23		23		
Siemens Clinitest		13		13		
Total Population	1	231	0	231	1	0
Flagging	***		***		***	***

2778 mIU/mL

<2.0 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
 Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.