



URINE MICROALBUMIN - CREATININE, qualitative

Microalbumin, semiquantative or qualitative

Name	Specimen 1										
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L	
Roche Chemstrip		2		2	9						
Siemens Clinitek Status					2					19	
Total Population	0	2	0	3	9	3	0	0	0	24	
Flagging	***		***	***			***				

Name	Specimen 2										
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L	
Roche Chemstrip		2		4	7						
Siemens Clinitek Status									9	12	
Total Population	0	2	0	4	7	1	0	0	10	17	
Flagging	***		***			***	***				

Creatinine, semiquantative

Name	Specimen 1					Specimen 2				
	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL
Siemens Clinitek Status		3	16	3					7	15
Total Population	0	3	21	5	0	0	1	0	8	20
Flagging	***				***	***	***			

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.