Bilirubin Supplemental, 2 vial

Name	Line No.	Specimen 1 Range & Type	Specimen 1 Mean	Specimen 1 SD	Specimen 2 Range & Type	Specimen 2 Mean	Specimen 2 SD	Specimen 3 Range & Type	Specimen 3 Mean	Specimen 3 SD	Specimen 4 Range & Type	Specimen 4 Mean	Specimen 4 SD	Specimen 5 Range & Type	Specimen 5 Mean	Specimen 5 SD	No. of Labs
Bilirubin, Direct																	
Initial Grouping by Reagent and Instrument																	
Beckman Olympus & Beck Olym AU 400/600/5400	1	4.8 - 7.1	P 5.95	0.48	1.3 - 2.1	C 1.71	0.28	-			-			-			11
Ortho Vitros & Ortho Vitros not DT or ECi	2	4.0 - 6.1	P 5.05	0.22	0.7 - 1.5	C 1.12	0.16	-			-			-			12
Siemens Dimension DBI & Siemens Dimension EXL	3	3.4 - 5.0	P 4.2	0.27	0.5 - 1.3	C 0.88	0.09	-			-			-			36
Initial Grouping by Reagent																	
Beckman Olympus	4	4.8 - 7.2	P 6.01	0.42	1.4 - 2.2	C 1.78	0.24	-			-			-			18
Ortho Vitros	5	4.0 - 6.0	P 5.0	0.2	0.7 - 1.5	C 1.08	0.15	-			-			-			20
Roche acid diazo	6	3.5 - 5.3	P 4.39	0.29	0.5 - 1.4	C 0.95	0.1	-			-			-			15
Siemens Dimension DBI	7	3.3 - 5.0	P 4.14	0.34	0.5 - 1.3	C 0.87	0.09	-			-			-			46
Initial Grouping by Sensitivity or Principle																	
Acid diazo methods	8	3.6 - 5.3	P 4.44	0.65	0.6 - 1.4	C 1.02	0.26	-			-			-			109
Diazonium ion methods	9	4.7 - 7.0	P 5.82	0.64	1.3 - 2.1	C 1.7	0.3	-			-			-			21
Total Population																	
Whole Population	10	3.7 - 5.6	P 4.66	0.81	0.7 - 1.5	C 1.13	0.37	-			-			-			133
Bilirubin, Neonatal																	
Initial Grouping by Reagent																	
Ortho Vitros	1	9.6 - 14.4	P 11.97	0.38	3.7 - 5.5	P 4.58	0.19	-			-			-			10
Initial Grouping by Sensitivity or Principle																	
Diazonium ion	2	9.6 - 14.4	P 11.97	0.38	3.7 - 5.5	P 4.58	0.19	-			-			-			10
Total Population																	
Whole Population	3	9.8 - 14.7	P 12.27	0.73	3.8 - 5.7	P 4.73	0.39	-			-			-			20
Bilirubin, Total supplemenal																	
Initial Grouping by Reagent and Instrument																	
Siemens Dimension & Siemens Dimension EXL	1	9.7 - 14.5	P 12.11	0.62	3.8 - 5.7	P 4.71	0.24	-			-			-			20
Initial Grouping by Reagent																	
Ortho Vitros	2	9.6 - 14.4	P 12.02	0.56	3.9 - 5.9	P 4.89	0.2	-			-			-			11
Siemens Dimension	3	9.7 - 14.5	P 12.1	0.71	3.8 - 5.7	P 4.71	0.28	-			-			-			26
Initial Grouping by Sensitivity or Principle																	
Diazonium ion	4	9.7 - 14.5	P 12.09	0.82	3.9 - 5.9	P 4.93	0.29	-			-			-			24
Diazo/caffeine-benzoate	5	9.7 - 14.6	P 12.14	0.71	3.8 - 5.7	P 4.74	0.29	-			-			-			30
Total Population																	
Whole Population	6	9.8 - 14.6	P 12.19	0.9	3.9 - 5.9	P 4.88	0.38	-			-			-			65