D-Dimer

Name	Line No.	Specimen 1 Range & Type	Specimen 1 Mean	Specimen 1 SD	Specimen 2 Range & Type	Specimen 2 Mean	Specimen 2 SD	Specimen 3 Range & Type	Specimen 3 Mean	Specimen 3 SD	Specimen 4 Range & Type	Specimen 4 Mean	Specimen 4 SD	Specimen 5 Range & Type	Specimen 5 Mean	Specimen 5 SD	No. of Labs
D-Dimer, quant ng/mL																	
Initial Grouping by Reagent																	
Alere Triage	1	636 - 1121	S 878.4	81.0	422 - 819	S 620.3	66.2	-			-			-			57
Initial Grouping by Sensitivity or Principle																	
low recovery	2	514 - 1186	S 849.6	112.0	315 - 877	S 595.7	93.7	-			-			-			64
Standardized method	3	596 - 1652	S 1123.9	175.9	346 - 749	S 547.6	67.2	-			-			-			11
Total Population																	
Whole Population	4	280 - 1574	S 926.6	215.6	192 - 1004	S 598.3	135.3	-			-			-			87
D-Dimer, quant ugFEU/mL																	
Initial Grouping by Reagent																	
Siemens Innovance	1	1.18 - 2.73	S 1.952	0.258	0.65 - 2.2	S 1.423	0.258	-			-			-			10
Initial Grouping by Sensitivity or Principle																	
Standardized method	2	0 - 3.42	S 1.701	0.573	0 - 2.52	S 1.204	0.439	-			-			-			20
Total Population																	
Whole Population	3	0 - 3.23	S 1.591	0.548	0 - 2.39	S 1.103	0.428	-			-			-			28