



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

THIRD QUADRIMESTER 2019

## PREGNANCY, waived

Name	Specimen 11			Specimen 12		
	Neg	Pos	Ind	Neg	Pos	Ind
Alere		6		4	2	
Beckman ICON 20/25 s/u		5		1	4	
CardinalHealth hCG Combo		22		5	17	
CLIAwaived		4		2	2	
Consult Diagnostics		31		6	25	
Fisher Sure-Vue hCG-STAT		8		3	5	
Fisher Sure-Vue urine hCG		24		6	18	
Henry Schein OneStep		6		2	2	2
McKesson		13		4	9	
Medline	1	7			8	
OneStep 3rd party, OEM waiv		10		1	9	
Quidel QuickVue		16		1	15	
Sekisui OSOM hCG-Urine	1	20		8	13	
Siemens Clinitest		16			16	
<b>Total Population</b>	<b>2</b>	<b>207</b>		<b>46</b>	<b>160</b>	<b>3</b>
<b>Flagging</b>	***			***		***
		55.0 mIU/mL			55.0 mIU/mL	

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.