



## URINE MICROALBUMIN - CREATININE, qualitative

**Microalbumin, semiquantitative or qualitative**

Name	Specimen 11									
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L
Roche Chemstrip		1		3	6	1				1
Siemens Clinitek Status			1						1	18
<b>Total Population</b>	<b>0</b>	<b>1</b>	<b>1</b>	<b>3</b>	<b>6</b>	<b>1</b>	<b>0</b>	<b>0</b>	<b>2</b>	<b>22</b>
<b>Flagging</b>	***						***	***		

Name	Specimen 12									
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L
Roche Chemstrip		1		1	9	1				
Siemens Clinitek Status						11				9
<b>Total Population</b>		<b>1</b>		<b>1</b>	<b>9</b>	<b>13</b>				<b>12</b>
<b>Flagging</b>				***						

**Creatinine, semiquantitative**

Name	Specimen 11					Specimen 12				
	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL
Siemens Clinitek Status		5	14	2			3	12	6	
<b>Total Population</b>	<b>0</b>	<b>6</b>	<b>17</b>	<b>2</b>	<b>1</b>	<b>0</b>	<b>3</b>	<b>14</b>	<b>8</b>	<b>1</b>
<b>Flagging</b>	***					***				***

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.