Urine Microalbumin - Creatinine

Name	Line No.	Specimen 1 Range & Type	Mean	SD	Specimen 2 Range & Type	Mean	SD	Specimen 3 Range & Type	Mean	SD	Specimen 4 Range & Type	Mean	SD	Specimen 5 Range & Type	Mean	SD	No. of Labs
Microalbumin, quantitative																	
Initial Grouping by Reagent																	
Beckman Olympus	1	66.6 - 111.1 P	88.86	20.62	100.9 - 168.2 P	134.58	24.37	-			-			-			24
Roche Tina-Quant	2	68.7 - 114.5 P	91.58	5.27	104.2 - 173.7 P	138.93	7.84	-			-			-			13
Siemens Dimension	3	68.6 - 114.4 P	91.51	2.71	100.5 - 167.5 P	133.97	16.4	-			-			-			23
Siemens DCA Vantage	4	69.6 - 116.0 P	92.78	5.39	109.7 - 182.9 P	146.32	8.03	-			-			-			18
Abbott Aeroset/Architect	5	58.1 - 96.8 P	77.47	34.22	88.6 - 147.6 P	118.09	52.06	-			-			-			11
Roche Cobas	6	69.5 - 115.9 P	92.7	3.69	107.7 - 179.6 P	143.65	10.43	-			-			-			11
Ortho Vitros	7	53.8 - 89.6 P	71.71	38.57	81.8 - 136.3 P	109.06	58.93	-			-			-			8
Initial Grouping by Sensitivity or Principle																	
Turbidimetric	8	66.2 - 110.4 P	88.28	18.18	99.7 - 166.1 P	132.88	27.05	-			-			-			80
Siemens turbidimetric	9	70.1 - 116.8 P	93.46	6.0	110.0 - 183.4 P	146.69	7.95	-			-			-			19
Chemiluminometric	10	58.1 - 96.8 P	77.47	34.22	88.6 - 147.6 P	118.09	52.06	-			-			-			11
All other methods	11	57.4 - 95.7 P	76.52	38.83	87.5 - 145.9 P	116.72	59.64	-			-			-			9
Turbidity immunoinhibition	12	70.0 - 116.6 P	93.3	2.01	107.9 - 179.8 P	143.85	2.17	-			-			-			4
Total Population																	
Whole Population	13	65.7 - 109.5 P	87.58	21.17	99.8 - 166.4 P	133.09	32.28	-			-			-			127
Creatinine, quantitative																	
Initial Grouping by Reagent																	
Beckman Coulter Synchron	1	59 - 79 P	69.0	6.1	59 - 79 P	69.0	5.9	-			-			-			5
Beckman Olympus	2	51 - 69 P	60.1	4.7	51 - 69 P	60.0	4.7	-			-			-			29
Synermed	3	62 - 84 P	73.0	4.0	62 - 83 P	72.5	3.7	-			-			-			22
Siemens Dimension	4	55 - 74 P	64.7	1.9	54 - 74 P	64.1	2.0	-			-			-			24
Siemens DCA Vantage	5	61 - 83 P	71.8	3.5	62 - 84 P	72.9	2.9	-			-			-			18
Roche Cobas	6	59 - 80 P	69.6	3.4	59 - 80 P	69.4	3.2	-			-			-			19
Abbott Aeroset/Architect	7	55 - 75 P	65.0	1.1	56 - 75 P	65.3	1.1	-			-			-			13
Ortho Vitros	8	56 - 75 P	65.6	3.0	55 - 75 P	65.0	2.1	-			-			-			7
Initial Grouping by Sensitivity or Principle																	
Jaffe/modified Jaffe	9	57 - 77 P	67.1	5.7	57 - 77 P	67.2	6.8	-			-			-			153
Total Population																	
Whole Population	10	57 - 77 P	67.2	5.9	57 - 77 P	67.1	6.9	-			-			-			157