



PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2011

ANTINUCLEAR ANTIBODY

Antinuclear Antibody Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	
Antibodies HEp-2		1	1			1	1		1		1
BioRad Microplate EIA HEp-2		7	7			7	7		7		7
BioRad Kallestat HEp-2		1	1			1	1		1		1
DiaMedix ELISA		4	4			4	4		4		4
Diasorin ANAFLUOR HEp-2		1	1			1	1		1		1
Immuno C HEp-2 Colorzyme		1	1			1	1		1		1
Immuno C HEp-2000 Colorzyme		4	4			4	4		4		4
Immuno C HEp-2000 IFA		2	2			2	2		2		2
Inova HEp-2 IgG		1	1			1	1		1		1
Inova QUANTA Lite ELISA		2	2			2	2		2		2
Inverness ANA Screen ELISA		5	4	1		5	4	1	4	1	5
Meridian ANA Screen		1	1			1	1		1		1
Scimedx Corp HEp-2		2	2			2	2		2		2
SeraQuest ANA Screening EIA		2	2			2	2		2		2
TheraTest EL-ANAScr		1	1			1	1		1		1
Zeus AtheNA		2	2			2	2		2		2
Zeus HEp-2		3	3			3	3		3		3
Total Population	0	40	39	1	0	40	39	1	39	1	40
Flagging	*****			*****	*****			*****		*****	

Anti-DNP / Anti-DNA Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	
Alpha Latex		13		13		13		13	13		13
ASI SLE latex agglutination		27		27		27		27	27		27
Diag Tech ANA-Check latex		21	1	20		21		21	20	1	21
Diagnostic Tech ANA-Check		8	1	7		8		8	8		8
Immunostics IMMUNO/LEX-sLE		2		2		2		2	2		2
Seradyn Seratest SLE latex		4		4		4		4	4		4
Total Population	0	76	2	74	0	76	0	76	74	2	76
Flagging	*****		*****		*****		*****		*****		*****

Anti-ENA Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	
GenBio ImmunoDOT EIA Panel		10	10			10	10		10		10
Seradyn Color Spot Panel 1		1	1			1	1		1		1
Total Population	0	11	11	0	0	11	11	0	11	0	11
Flagging	*****			*****	*****			*****		*****	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.