



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2011**

**C-REACTIVE PROTEIN**

C-REACTIVE PROTEIN	Specimen 1		Specimen 2		No. of Labs
Method	NEG	POS	NEG	POS	
ASI		19	2	17	19
Biokit Rhe-jet		1		1	1
Fisher Sure-Vue		6	1	5	6
Immunostics immuno/lex-crp		11		11	11
Randox CRP Latex		2	1	1	2
Seradyn SeraTest Latex		2		2	2
Siemens RapiTex		2		2	2
Stanbio #1130		2		2	2
Teco Diag latex slide test		2		2	2
True-Medix Trutest CRP		1		1	1
<b>Total Population</b>	<b>1</b>	<b>106</b>	<b>10</b>	<b>97</b>	<b>107</b>
<b>Flagging</b>	<b>*****</b>		<b>*****</b>		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.