



PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2011

HELICOBACTER PYLORI

H. pylori

Method	Specimen 1		Specimen 2		No. of Labs
	Neg	Pos	Neg	Pos	
ABI SureStep		4	4		4
Alfa Instant-View		15	15		15
American Diag Prod Rapid		7	7		7
Beckman Coulter Icon	1	14	15		15
BioRad GAP IgG ELISA		2	2		2
BioStar Acceava		3	3		3
BioUSA		6	6		6
Cardinal Health SP	1	11	9	3	12
Consult	1	6	7		7
Diag Test Group Clarity		4	4		4
DiaMedix EIA		2	2		2
Fisher Sure-Vue IgG	1	27	26	2	28
Genzyme OSOM		2	2		2
Germaine		5	4	1	5
Henry Schein One Step		7	6	1	7
Innovacon	1	1	2		2
Instant Technologies iScrn		4	4		4
Inverness Clearview I/II	1	36	35	2	37
Inverness IgG ELISA		5	5		5
Jant Pharmacal Accutest WB		2	2		2
LifeSign Status		2	2		2
McKesson Medi-Lab		8	8		8
Meridian ImmunoCard IgG EIA	1	6	6	1	7
Micro Detect IgG		2	2		2
Polymedco Poly stat IgG	1	15	16		16
Quidel QuickVue gII	1	81	81	1	82
Select Diagnostics		4	4		4
Siemens Immulite 2000		10	10		10
SP Rapid		7	6	1	7
Stanbio	1	1	2		2
Thermo/Inverness Acceava		2	2		2
Total Population	11	309	307	13	320
Flagging	*****			*****	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.