



PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2011

RUBELLA

Rubella IgG Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Abbott AxSYM IgG only	10			10		10		10		10	10
ASI Rubella Test	32			32		32	1	31		32	32
Beckman Access IgG	12			12		12		12		12	12
Biokit Rbgen/Fisher SureVue	19			19	1	18		19		19	19
Biokit/SV Select Rubella	3			3		3		3		3	3
bioMerieux Vidas IgG ELFA	8			8		8		8		8	8
BioRad Rubella IgG EIA	3			3		3		3		3	3
Diamedix Microassay IgG	6			6		6		6		6	6
DiaSorin Rubella IgG ELISA	2			2		2		2		2	2
Inverness IgG ELISA	2			2		2		2		2	2
Inverness Rubella-Plus	11		1	10	1	10	1	10		11	11
J&J Vitros	8			8		8		8		8	8
Remel Ruba-Test	3			3		3		3		3	3
Roche Cobas/Elecsys Rub IgG	5			5		5		5		5	5
Seradyn Color Slide Rubella	2			2		2		2		2	2
SeraQuest Rubella IgG ELISA	2			2		2		2		2	2
Siemens Immulite 2000, 2500	11			11		11		11		11	11
Siemens Immulite IgG	7			7		7		7		7	7
Siemens Immuno 1/Advia IgG	4	2		6		6		6		6	6
True Medix Trutest Rubella	2			2		2		2		2	2
Wampole Impact	4			4		4		4		4	4
Zeus ELISA	2			2		2		2		2	2
Total Population	166	2	1	167	2	166	2	166	0	168	168
Flagging		*****	*****		*****		*****		*****		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.