



PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2011

STREP GROUP A ANTIGEN SCREEN

Strep A Screen Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Abbott Signify	6			6	6		6			6	6
BD Chek	2			2	2		2			2	2
BD Directigen EZ	6			6	6		6			6	6
Beckman Coulter ICON ser	6			6	6		6			6	6
Cardinal Health SP	21			21	21		21			21	21
Cardinal Health SP nonwvd	5			5	5		5			5	5
Consult Twist	4			4	4		4			4	4
Diag Test Group Clarity	2			2	2		2			2	2
Fish Sure-Vue Signature, wv	6			6	6		6			6	6
Fisher Sure-Vue, nonwaived	12			12	12		12			12	12
Fisher Sure-Vue, waived	44			44	44		44			44	44
Genzyme OSOM	31			31	30	1	29	2	1	30	31
Genzyme OSOM Ultra	4			4	4		4			4	4
Genzyme OSOM Ultra	50			50	50		50			50	50
Genzyme OSOM, waived	17			17	17		17			17	17
Genzyme Rapid Strep A	4			4	4		4			4	4
Germaine StrepAim	3			3	3		3			3	3
Henry Schein OneStep	13			13	13		13			13	13
Henry Schein OneStep+, Wv	10		1	9	10		10		1	9	10
Immunostics Detector	3			3	3		3			3	3
Immunostics immune/Strep A	3			3	3		3			3	3
Invern/BioStar Acceava	78			78	78		78			78	78
Inverness Binax Now	3			3	3		3			3	3
Inverness Clearview	21			21	21		21			21	21
Inverness Signify	5			5	5		5			5	5
McKesson Twist	15			15	15		15			15	15
McKesson, nonwaived	7			7	7		7			7	7
McKesson, waived	6			6	5	1	6			6	6
Meridian Diag ImmunoCard	9			9	9		9			9	9
NDC Pro Advantage	3			3	3		3			3	3
PEP	1			1	1		1			1	1
Polymedco Poly Stat	5			5	5		5			5	5
Polymedco PolyStat nonwaivd	2			2	2		2			2	2
PSS Select / Consult Diag	18			18	18		18			18	18
Quidel QuickVue +	7			7	7		7			7	7
Quidel QuickVue +	63			63	63		62	1	1	61	63
Quidel QuickVue Dipstick	40			40	39		38	1	1	39	40
Rediscreen	1			1	1		1			1	1
SA Scientific Strep Alert	3			3	3		3			3	3
Select Diagnostics	7			7	7		7			7	7
Stanbio QuSTICK	2			2	2		2			2	2
Total Population	680	4	13	671	677	5	674	9	14	668	684
Flagging		*****	*****			*****		*****	*****		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.