



PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2011

ToRCH

Cytomegalovirus

Method
DiaSorin LIAISON
Total Population
Flagging

Specimen 1	
Neg	Pos
	1
0	2

Specimen 2	
Neg	Pos
	1
0	2

Specimen 3	
Neg	Pos
1	
2	0

Herpes Type II or I/II, IgG

Method
DiaSorin LIAISON
Trinity Biotech Captia
Total Population
Flagging

Specimen 1		
Neg	Pos for I or II	Pos Type II
3		
1		
5	1	0

Specimen 2		
Neg	Pos for I or II	Pos Type II
		3
1		
2	1	3

Specimen 3		
Neg	Pos for I or II	Pos Type II
3		
1		
4	1	1

Due to a lack of participant consensus, specimen 2 was not evaluated this event.

Herpes Type I, IgG or non-spec

Method
DiaSorin LIAISON
Total Population
Flagging

Specimen 1	
Neg	Pos
	3
0	4

Specimen 2	
Neg	Pos
	3
0	4

Specimen 3	
Neg	Pos
	3
0	4

Herpes Type II or I/II, IgM

Method
Total Population
Flagging

Specimen 1	
Neg	Pos IgG NonSpec
2	0

Specimen 2	
Neg	Pos IgG NonSpec
1	1

Specimen 3	
Neg	Pos IgG NonSpec
1	1

Due to a lack of participant consensus, specimen 3 was not evaluated this event.

Herpes Type I, IgM

Method
Total Population
Flagging

Specimen 1	
Neg	Pos
0	3

Specimen 2	
Neg	Pos
3	0

Specimen 3	
Neg	Pos
3	0

Rubella IgM

Method
Bio-Rad Rubella IgM
Diamedix IgM Capture
Inverness IgM ELISA
Siemens Immulite 2000, 2500
Total Population
Flagging

Specimen 1	
Neg	Pos
	1
	1
	1
	1
0	4

Specimen 2	
Neg	Pos
1	
1	
1	
1	
4	0

Specimen 3	
Neg	Pos
1	
1	
1	
1	
4	0

Toxoplasma

Method
DiaSorin LIAISON
Total Population
Flagging

Specimen 1	
Neg	Pos
	2
0	3

Specimen 2	
Neg	Pos
2	
3	0

Specimen 3	
Neg	Pos
2	
3	0

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.