



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2016

Infectious Mononucleosis Waived

Infectious Mono Waived

Method
Alere Acceava
Alere Clearview
Consult Diagnostics
Quidel QuickVue+
Sekisui OSOM Mono Whole Bl
Total Population
Flagging

Specimen 1

Neg	Pos
9	
9	
7	
6	
11	
57	0

Specimen 2

Neg	Pos
	9
	9
	7
	6
	11
0	57

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.