



**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2016**

**RUBELLA**

Rubella IgG Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere Rubella-Plus	7			7		7	7			7	7
ASI Rubella Test	20			20		20	20			20	20
Beckman Access IgG	9			9		9	9			9	9
Biokit Rbgen/Fisher SureVue	10			10		10	10			10	10
Biokit/SV Select Rubella	6			6		6	6			6	6
bioMerieux Vidas IgG ELFA	4			4		4	4			4	4
Roche Elecsys	8			8		8	8			8	8
Siemens Immulite 2000, 2500	6			6		6	6			6	6
Siemens Immulite IgG	5		1	4		5	4	1		5	5
<b>Total Population</b>	<b>105</b>	<b>1</b>	<b>1</b>	<b>106</b>	<b>1</b>	<b>106</b>	<b>104</b>	<b>2</b>	<b>106</b>	<b>0</b>	<b>106</b>
<b>Flagging</b>		***	***		***			***		***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.