



PROFICIENCY TESTING SERVICE  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTIC PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2016**

**Sickle Screen**

Method
AMSCO
ASI Sickle Cell Test
Chembio Sickle-Stat
Pac Hemostasis SickleScreen
Siemens Sickle-Sol Test
Streck Sickle-Chex Solub
Streck Sickledex
<b>Total Population</b>
<b>Flagging</b>

Specimen 1	
Neg	Pos
2	
5	
1	
5	
1	
3	
2	1
<b>19</b>	<b>1</b>
	<b>***</b>

Specimen 2	
Neg	Pos
	2
1	4
	1
	5
	1
	3
1	2
<b>2</b>	<b>18</b>
<b>***</b>	

No. of Labs
2
5
1
5
1
3
3
3
<b>20</b>

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.