



## HEMATOLOGY WITH DIFF E

Name	Line No.	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	
<b>Leukocytes - Module E</b>																	
<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	18.1 - 24.6 P 21.35	0.44	6.6 - 8.9 P 7.75	0.17	3.4 - 4.6 P 3.99	0.09	3.4 - 4.6 P 3.98	0.08	6.6 - 8.9 P 7.73	0.15	58					
Sysmex XN-1000	2	17.1 - 23.1 P 20.11	0.45	6.1 - 8.3 P 7.2	0.23	3.3 - 4.4 P 3.85	0.08	3.3 - 4.4 P 3.86	0.08	6.1 - 8.3 P 7.18	0.19	10					
Sysmex XT 4000i	3	17.2 - 23.2 P 20.21	0.7	6.4 - 8.6 P 7.49	0.21	3.3 - 4.5 P 3.89	0.08	3.3 - 4.4 P 3.84	0.16	6.4 - 8.6 P 7.5	0.23	10					
Sysmex XE-2100	4	17.7 - 23.9 P 20.77	0.95	6.5 - 8.9 P 7.7	0.22	3.3 - 4.4 P 3.83	0.25	3.4 - 4.6 P 4.0	0.14	6.5 - 8.9 P 7.7	0.37	4					
Sysmex XT 1800i	5	16.7 - 22.6 P 19.67	0.76	6.5 - 8.7 P 7.6	0.14	3.3 - 4.4 P 3.83	0.14	3.3 - 4.5 P 3.88	0.13	6.4 - 8.6 P 7.52	0.18	6					
Sysmex XT 2000i	6	17.3 - 23.3 P 20.3	0.6	6.5 - 8.7 P 7.59	0.23	3.3 - 4.4 P 3.83	0.21	3.4 - 4.5 P 3.95	0.17	6.4 - 8.7 P 7.53	0.11	11					
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	18.1 - 24.6 P 21.35	0.44	6.6 - 8.9 P 7.75	0.17	3.4 - 4.6 P 3.99	0.09	3.4 - 4.6 P 3.98	0.08	6.6 - 8.9 P 7.73	0.15	58					
Sysmex XE/XT/XN series	8	17.2 - 23.2 P 20.19	0.72	6.4 - 8.6 P 7.49	0.28	3.3 - 4.4 P 3.86	0.16	3.3 - 4.5 P 3.89	0.15	6.3 - 8.6 P 7.45	0.27	44					
<b>Total Population</b>																	
Whole Population	9	17.7 - 24.0 P 20.84	0.82	6.5 - 8.8 P 7.64	0.26	3.3 - 4.5 P 3.93	0.14	3.3 - 4.5 P 3.94	0.13	6.5 - 8.8 P 7.61	0.25	103					
<b>Erythrocytes - Module E</b>																	
<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	5.31 - 5.99 P 5.652	0.055	4.34 - 4.89 P 4.616	0.045	2.17 - 2.45 P 2.308	0.028	2.17 - 2.45 P 2.312	0.03	4.33 - 4.88 P 4.608	0.047	58					
Sysmex XN-1000	2	5.31 - 5.99 P 5.647	0.063	4.35 - 4.91 P 4.632	0.048	2.17 - 2.44 P 2.305	0.025	2.15 - 2.43 P 2.292	0.034	4.32 - 4.87 P 4.599	0.052	10					
Sysmex XT 4000i	3	5.36 - 6.04 P 5.702	0.093	4.43 - 5.0 P 4.718	0.04	2.26 - 2.55 P 2.403	0.021	2.25 - 2.54 P 2.397	0.031	4.44 - 5.0 P 4.719	0.039	10					
Sysmex XE-2100	4	5.3 - 5.97 P 5.633	0.012	4.39 - 4.95 P 4.67	0.029	2.22 - 2.5 P 2.363	0.024	2.23 - 2.51 P 2.367	0.019	4.38 - 4.94 P 4.657	0.052	4					
Sysmex XT 1800i	5	5.32 - 6.0 P 5.662	0.062	4.4 - 4.96 P 4.682	0.046	2.24 - 2.52 P 2.382	0.032	2.23 - 2.51 P 2.372	0.034	4.4 - 4.96 P 4.683	0.066	6					
Sysmex XT 2000i	6	5.39 - 6.08 P 5.737	0.087	4.42 - 4.98 P 4.7	0.056	2.24 - 2.53 P 2.386	0.038	2.25 - 2.53 P 2.389	0.028	4.42 - 4.99 P 4.707	0.071	11					
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	5.31 - 5.99 P 5.652	0.055	4.34 - 4.89 P 4.616	0.045	2.17 - 2.45 P 2.308	0.028	2.17 - 2.45 P 2.312	0.03	4.33 - 4.88 P 4.608	0.047	58					
Sysmex XE/XT series	8	5.34 - 6.03 P 5.686	0.083	4.4 - 4.96 P 4.683	0.055	2.23 - 2.51 P 2.37	0.047	2.22 - 2.51 P 2.364	0.051	4.35 - 4.91 P 4.632	0.301	44					
<b>Total Population</b>																	
Whole Population	9	5.33 - 6.01 P 5.667	0.071	4.37 - 4.92 P 4.645	0.06	2.19 - 2.48 P 2.335	0.048	2.19 - 2.48 P 2.335	0.048	4.34 - 4.9 P 4.618	0.202	103					
<b>Hemoglobin - Module E</b>																	
<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	16.3 - 18.7 P 17.5	0.17	12.2 - 14.1 P 13.17	0.15	5.3 - 6.1 P 5.67	0.08	5.3 - 6.1 P 5.66	0.07	12.2 - 14.1 P 13.17	0.12	58					
Sysmex XN-1000	2	16.2 - 18.7 P 17.47	0.17	12.3 - 14.1 P 13.22	0.16	5.3 - 6.1 P 5.71	0.08	5.3 - 6.1 P 5.73	0.09	12.3 - 14.1 P 13.2	0.15	10					
Sysmex XT 4000i	3	16.1 - 18.6 P 17.36	0.2	12.3 - 14.1 P 13.21	0.07	5.4 - 6.2 P 5.77	0.06	5.4 - 6.2 P 5.77	0.05	12.3 - 14.1 P 13.22	0.1	10					
Sysmex XE-2100	4	16.5 - 19.0 P 17.73	0.09	12.4 - 14.3 P 13.33	0.05	5.5 - 6.3 P 5.87	0.09	5.5 - 6.3 P 5.9	0.08	12.5 - 14.4 P 13.43	0.12	4					
Sysmex XT 1800i	5	16.1 - 18.5 P 17.27	0.22	12.2 - 14.1 P 13.17	0.18	5.3 - 6.2 P 5.75	0.08	5.3 - 6.2 P 5.75	0.05	12.3 - 14.1 P 13.2	0.12	6					
Sysmex XT 2000i	6	16.3 - 18.7 P 17.48	0.36	12.3 - 14.2 P 13.25	0.16	5.3 - 6.2 P 5.75	0.08	5.3 - 6.1 P 5.73	0.1	12.3 - 14.2 P 13.25	0.16	11					
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	16.3 - 18.7 P 17.5	0.17	12.2 - 14.1 P 13.17	0.15	5.3 - 6.1 P 5.67	0.08	5.3 - 6.1 P 5.66	0.07	12.2 - 14.1 P 13.17	0.12	58					
Sysmex XE/XT series	8	16.2 - 18.7 P 17.43	0.29	12.3 - 14.2 P 13.23	0.15	5.4 - 6.2 P 5.76	0.09	5.4 - 6.2 P 5.76	0.1	12.3 - 14.2 P 13.24	0.15	44					
<b>Total Population</b>																	
Whole Population	9	16.2 - 18.7 P 17.47	0.23	12.3 - 14.1 P 13.2	0.15	5.3 - 6.1 P 5.71	0.09	5.3 - 6.1 P 5.7	0.1	12.3 - 14.1 P 13.2	0.14	103					
<b>Hematocrit - Module E</b>																	
<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	48.9 - 55.1 P 51.97	0.89	36.7 - 41.4 P 39.08	0.54	16.3 - 18.4 P 17.38	0.24	16.3 - 18.4 P 17.39	0.26	36.6 - 41.3 P 38.98	0.57	58					
Sysmex XT 4000i	2	49.0 - 55.2 P 52.08	1.22	37.1 - 41.9 P 39.5	0.68	16.9 - 19.0 P 17.93	0.31	16.8 - 18.9 P 17.87	0.29	37.1 - 41.9 P 39.5	0.72	10					
Sysmex XN-1000	3	48.0 - 54.2 P 51.09	0.78	36.4 - 41.0 P 38.7	0.56	16.0 - 18.0 P 16.98	0.29	15.9 - 17.9 P 16.88	0.33	36.1 - 40.7 P 38.44	0.56	10					
Sysmex XE-2100	4	47.8 - 53.8 P 50.8	0.29	36.5 - 41.1 P 38.8	0.37	16.2 - 18.3 P 17.27	0.19	16.2 - 18.3 P 17.27	0.19	36.4 - 41.0 P 38.67	0.6	4					
Sysmex XT 1800i	5	49.1 - 55.3 P 52.2	0.87	37.2 - 41.9 P 39.53	0.55	16.8 - 19.0 P 17.88	0.23	16.7 - 18.9 P 17.8	0.32	37.1 - 41.9 P 39.52	0.69	6					

Sysmex XT 2000i	6	49.1 - 55.4	P 52.22	0.72	36.9 - 41.6	P 39.27	0.4	16.6 - 18.8	P 17.71	0.21	16.7 - 18.8	P 17.75	0.17	37.0 - 41.7	P 39.36	0.48	11
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	48.9 - 55.1	P 51.97	0.89	36.7 - 41.4	P 39.08	0.54	16.3 - 18.4	P 17.38	0.24	16.3 - 18.4	P 17.39	0.26	36.6 - 41.3	P 38.98	0.57	58
Sysmex XE/XT series	8	48.7 - 54.9	P 51.81	1.02	36.9 - 41.6	P 39.22	0.65	16.5 - 18.6	P 17.59	0.46	16.5 - 18.6	P 17.55	0.48	36.8 - 41.5	P 39.19	0.77	44
<b>Total Population</b>																	
Whole Population	9	48.8 - 55.0	P 51.91	0.95	36.8 - 41.5	P 39.14	0.59	16.4 - 18.5	P 17.47	0.37	16.4 - 18.5	P 17.46	0.38	36.7 - 41.4	P 39.07	0.67	103

### Platelets - Module E

<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	309 - 515	P 411.9	12.8	157 - 261	P 209.2	7.8	37 - 62	P 49.8	2.5	37 - 62	P 49.5	3.4	156 - 261	P 208.4	9.3	58
Sysmex XT 4000i	2	317 - 528	P 422.0	23.9	164 - 273	P 218.2	8.6	41 - 69	P 55.0	3.0	40 - 66	P 53.1	3.0	160 - 266	P 212.9	9.0	10
Sysmex XN-1000	3	322 - 537	P 429.2	12.1	162 - 269	P 215.4	7.1	35 - 58	P 46.7	3.3	35 - 58	P 46.7	2.7	160 - 267	P 213.3	7.3	10
Sysmex XE-2100	4	291 - 485	P 388.0	2.1	148 - 247	P 197.5	5.0	33 - 55	P 44.3	2.5	34 - 57	P 45.8	2.9	148 - 246	P 196.8	6.9	4
Sysmex XT 1800i	5	306 - 510	P 407.8	18.1	158 - 263	P 210.0	9.3	39 - 64	P 51.5	2.1	40 - 66	P 52.7	3.9	158 - 264	P 211.0	6.9	6
Sysmex XT 2000i	6	325 - 542	P 433.5	12.7	163 - 271	P 217.0	4.1	41 - 69	P 55.3	2.7	40 - 67	P 53.7	2.6	164 - 273	P 218.3	6.0	11
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	309 - 515	P 411.9	12.8	157 - 261	P 209.2	7.8	37 - 62	P 49.8	2.5	37 - 62	P 49.5	3.4	156 - 261	P 208.4	9.3	58
Sysmex XE/XT series	8	316 - 527	P 421.7	22.6	161 - 268	P 214.2	10.0	39 - 64	P 51.5	5.2	38 - 64	P 50.9	4.4	160 - 266	P 212.8	9.7	44
<b>Total Population</b>																	
Whole Population	9	312 - 520	P 416.3	18.4	159 - 264	P 211.4	9.1	38 - 63	P 50.5	4.0	38 - 63	P 50.1	3.9	158 - 263	P 210.4	9.7	103

### Neutrophil - Module E

<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	49.9 - 56.5	S 53.21	1.11	43.1 - 53.0	S 48.03	1.64	54.0 - 62.3	S 58.14	1.39	52.1 - 64.0	S 58.02	1.99	43.4 - 53.0	S 48.2	1.61	58
Sysmex XT 4000i	2	59.9 - 63.8	S 61.88	0.64	53.0 - 57.6	S 55.29	0.77	63.6 - 70.1	S 66.84	1.08	64.1 - 69.3	S 66.7	0.88	53.5 - 57.1	S 55.3	0.61	9
Sysmex XN-1000	3	44.9 - 76.3	S 60.59	5.23	43.1 - 61.6	S 52.34	3.08	53.5 - 72.9	S 63.19	3.23	54.1 - 72.8	S 63.46	3.11	41.2 - 64.8	S 53.0	3.92	10
Sysmex XE-2100	4	53.8 - 66.2	S 60.0	2.05	53.8 - 54.1	S 53.97	0.05	61.3 - 68.2	S 64.77	1.15	55.0 - 73.0	S 64.03	3.0	51.4 - 57.8	S 54.63	1.07	4
Sysmex XT 1800i	5	58.8 - 64.1	S 61.43	0.88	52.7 - 57.5	S 55.07	0.8	62.7 - 70.1	S 66.4	1.24	62.2 - 71.4	S 66.83	1.53	52.9 - 57.1	S 55.0	0.69	6
Sysmex XT 2000i	6	57.2 - 66.4	S 61.82	1.54	52.3 - 57.6	S 54.95	0.89	58.7 - 73.0	S 65.81	2.38	62.0 - 71.2	S 66.58	1.53	53.0 - 57.3	S 55.15	0.72	11
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	49.9 - 56.5	S 53.21	1.11	43.1 - 53.0	S 48.03	1.64	54.0 - 62.3	S 58.14	1.39	52.1 - 64.0	S 58.02	1.99	43.4 - 53.0	S 48.2	1.61	55
Sysmex XE/XT series	8	52.5 - 70.0	S 61.24	2.93	47.9 - 60.5	S 54.21	2.09	57.6 - 73.1	S 65.37	2.58	57.8 - 73.4	S 65.64	2.6	47.6 - 61.4	S 54.47	2.3	43
<b>Total Population</b>																	
Whole Population	9	42.6 - 72.1	S 57.32	4.91	40.0 - 61.5	S 50.73	3.58	49.0 - 73.7	S 61.34	4.11	48.1 - 74.6	S 61.37	4.42	39.9 - 62.0	S 50.93	3.68	102

### Lymphocyte % - Module E

<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	17.0 - 27.6	S 22.31	1.76	19.7 - 39.7	S 29.69	3.33	10.7 - 20.7	S 15.7	1.66	11.1 - 20.4	S 15.73	1.56	21.1 - 39.1	S 30.09	3.0	58
Sysmex XT 4000i	2	18.7 - 25.1	S 21.91	1.07	26.5 - 31.4	S 28.96	0.82	12.4 - 18.1	S 15.23	0.95	12.8 - 17.6	S 15.18	0.8	26.6 - 30.8	S 28.72	0.71	9
Sysmex XN-1000	3	15.6 - 26.0	S 20.8	1.73	14.5 - 40.8	S 27.65	4.39	8.2 - 22.7	S 15.44	2.42	7.6 - 22.4	S 14.97	2.47	14.9 - 39.3	S 27.1	4.08	10
Sysmex XE-2100	4	16.7 - 30.1	S 23.37	2.24	29.4 - 31.1	S 30.27	0.29	12.5 - 21.5	S 17.0	1.49	11.2 - 25.9	S 18.57	2.44	28.2 - 32.1	S 30.17	0.65	4
Sysmex XT 1800i	5	19.8 - 24.3	S 22.02	0.75	26.5 - 31.8	S 29.15	0.88	13.4 - 18.5	S 15.98	0.85	12.6 - 17.8	S 15.22	0.86	28.2 - 30.6	S 29.4	0.4	6
Sysmex XT 2000i	6	20.3 - 22.6	S 21.46	0.38	27.2 - 31.4	S 29.26	0.7	12.2 - 19.6	S 15.88	1.24	12.0 - 18.0	S 15.02	0.99	26.6 - 32.3	S 29.43	0.96	11
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	17.0 - 27.6	S 22.31	1.76	19.7 - 39.7	S 29.69	3.33	10.7 - 20.7	S 15.7	1.66	11.1 - 20.4	S 15.73	1.56	21.1 - 39.1	S 30.09	3.0	58
Sysmex XE/XT series	8	16.7 - 26.9	S 21.79	1.69	21.1 - 37.0	S 29.08	2.64	10.4 - 21.3	S 15.87	1.81	9.3 - 21.7	S 15.47	2.06	21.1 - 36.8	S 28.93	2.62	43
<b>Total Population</b>																	
Whole Population	9	16.9 - 27.3	S 22.1	1.75	20.3 - 38.7	S 29.46	3.07	10.6 - 20.9	S 15.78	1.72	10.3 - 21.0	S 15.62	1.79	20.9 - 38.3	S 29.61	2.89	102

### Monocyte % - Module E

<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	0 - 5.7	S 1.59	1.36	0 - 8.7	S 2.27	2.13	0 - 7.0	S 2.45	1.51	0 - 6.0	S 2.21	1.27	0 - 8.9	S 2.2	2.24	58
Sysmex XT 4000i	2	1.4 - 3.4	C 2.38	0.2	1.9 - 3.9	C 2.93	0.23	0.4 - 4.9	S 2.66	0.75	0.8 - 5.2	S 2.98	0.74	1.8 - 4.2	S 2.99	0.39	9
Sysmex XN-1000	3	0 - 6.0	S 2.24	1.26	0 - 10.1	S 2.88	2.42	0 - 8.9	S 2.35	2.2	0 - 9.8	S 2.65	2.4	0 - 6.8	S 2.55	1.41	10
Sysmex XE-2100	4	0.7 - 4.2	S 2.43	0.57	0.8 - 4.5	S 2.67	0.61	1.1 - 4.5	S 2.8	0.57	0.4 - 4.9	S 2.63	0.76	1.2 - 3.4	S 2.3	0.36	4
Sysmex XT 1800i	5	1.1 - 4.2	S 2.62	0.51	1.5 - 4.1	S 2.8	0.44	0.6 - 4.2	S 2.43	0.6	0 - 6.1	S 2.85	1.09	0.9 - 4.7	S 2.78	0.63	6
Sysmex XT 2000i	6	0.6 - 3.4	S 2.0	0.47	1.7 - 3.7	C 2.7	0.25	0.5 - 4.3	S 2.36	0.64	0 - 6.0	S 2.72	1.09	1.5 - 3.7	S 2.61	0.38	11
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	0 - 5.7	S 1.59	1.36	0 - 8.7	S 2.27	2.13	0 - 7.0	S 2.45	1.51	0 - 6.0	S 2.21	1.27	0 - 8.9	S 2.2	2.24	58
Sysmex XE/XT series	8	0 - 4.5	S 2.25	0.75	0 - 6.6	S 2.76	1.28	0 - 6.1	S 2.43	1.23	0 - 7.1	S 2.74	1.46	0.2 - 5.1	S 2.65	0.83	43

<b>Total Population</b>																	
Whole Population	9	0 - 5.4	S 1.86	1.19	0 - 8.0	S 2.47	1.84	0 - 6.6	S 2.43	1.4	0 - 6.6	S 2.45	1.38	0 - 7.8	S 2.37	1.81	102

### Eosinophil % - Module E

<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	10.2 - 17.5	S 13.86	1.21	7.8 - 19.3	S 13.54	1.91	7.6 - 23.9	S 15.79	2.71	8.4 - 23.0	S 15.72	2.44	7.4 - 19.5	S 13.45	2.01	58
Sysmex XT 4000i	2	13.0 - 15.0	C 13.99	0.28	11.8 - 13.8	C 12.83	0.3	13.2 - 17.2	S 15.19	0.68	14.1 - 16.1	C 15.13	0.31	12.0 - 14.0	C 12.98	0.29	9
Sysmex XN-1000	3	0.4 - 23.8	S 12.1	3.91	10.6 - 16.0	S 13.28	0.91	11.0 - 19.5	S 15.25	1.41	10.0 - 20.3	S 15.14	1.72	10.8 - 15.9	S 13.37	0.85	10
Sysmex XE-2100	4	12.3 - 16.1	S 14.2	0.65	11.6 - 14.6	S 13.1	0.5	13.4 - 17.4	S 15.43	0.66	13.8 - 15.8	C 14.77	0.05	11.9 - 13.9	C 12.9	0.14	4
Sysmex XT 1800i	5	12.8 - 15.0	S 13.93	0.37	11.7 - 14.2	S 12.98	0.41	14.0 - 16.3	S 15.18	0.39	13.9 - 16.3	S 15.1	0.39	11.8 - 13.8	C 12.82	0.27	6
Sysmex XT 2000i	6	12.8 - 15.3	S 14.03	0.42	11.9 - 13.9	C 12.9	0.24	13.4 - 16.4	S 14.92	0.5	14.0 - 16.5	S 15.23	0.42	11.8 - 13.8	C 12.75	0.31	11
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	10.2 - 17.5	S 13.86	1.21	7.8 - 19.3	S 13.54	1.91	7.6 - 23.9	S 15.79	2.71	8.4 - 23.0	S 15.72	2.44	7.4 - 19.5	S 13.45	2.01	55
Sysmex XE/XT series	8	7.2 - 19.9	S 13.56	2.11	11.3 - 14.7	S 12.98	0.56	12.5 - 17.8	S 15.14	0.88	12.3 - 17.9	S 15.14	0.93	11.3 - 14.6	S 12.96	0.56	43
<b>Total Population</b>																	
Whole Population	9	8.7 - 18.7	S 13.73	1.67	8.8 - 17.8	S 13.28	1.5	9.1 - 21.9	S 15.5	2.12	9.6 - 21.3	S 15.45	1.94	8.6 - 17.9	S 13.24	1.56	102

### Basophil % - Module E

<b>Initial Grouping by Reagent</b>																	
Sysmex XS-800i, 1000i	1	0 - 16.9	S 7.09	3.27	0 - 14.2	S 6.07	2.72	0 - 17.3	S 7.23	3.36	0 - 17.7	S 7.33	3.45	0 - 13.9	S 5.96	2.66	58
Sysmex XT 4000i	2	78.8 - 82.7	S 80.75	0.66	66.5 - 70.9	S 68.66	0.74	81.4 - 85.3	S 83.32	0.65	80.0 - 88.0	S 83.99	1.32	67.0 - 70.4	S 68.72	0.57	9
Sysmex XN-1000	3	0 - 9.8	S 3.57	2.08	0 - 9.2	S 3.54	1.89	0 - 9.7	S 3.41	2.11	0 - 9.9	S 3.36	2.18	0 - 9.2	S 3.58	1.87	10
Sysmex XE-2100	4	76.2 - 80.2	S 78.2	0.67	65.9 - 72.6	S 69.23	1.11	73.8 - 98.1	S 85.97	4.05	79.2 - 86.4	S 82.8	1.19	67.1 - 70.3	S 68.7	0.54	4
Sysmex XT 1800i	5	79.5 - 82.5	S 81.0	0.49	67.3 - 70.6	S 68.95	0.56	81.6 - 84.9	S 83.28	0.56	81.8 - 85.8	S 83.8	0.68	64.8 - 71.7	S 68.27	1.14	6
Sysmex XT 2000i	6	77.9 - 84.2	S 81.07	1.06	66.2 - 71.4	S 68.79	0.86	82.6 - 85.5	S 84.06	0.49	81.4 - 86.1	S 83.72	0.79	66.1 - 71.5	S 68.8	0.89	11
<b>Initial Grouping by Sensitivity or Principle</b>																	
Sysmex XS series	7	0 - 16.6	S 6.56	3.36	0 - 14.0	S 5.69	2.76	0 - 17.1	S 6.66	3.48	0 - 17.5	S 6.73	3.58	0 - 13.7	S 5.6	2.69	68
Sysmex XE/XT series	8	76.5 - 84.4	S 80.42	1.32	66.4 - 71.3	S 68.85	0.8	79.1 - 88.7	S 83.9	1.61	80.7 - 86.8	S 83.75	1.03	66.1 - 71.2	S 68.62	0.85	33
<b>Total Population</b>																	
Whole Population	9	0 - 131.6	S 29.19	34.15	0 - 112.6	S 25.03	29.2	0 - 137.5	S 30.32	35.71	0 - 137.2	S 30.33	35.61	0 - 112.3	S 24.91	29.12	102