



**PROFICIENCY TESTING SERVICE  
AMERICAN ASSOCIATION OF BIOANALYSTS**  
205 West Levee St. ♦ Brownsville, TX 78520-5596  
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2018**

**CHLAMYDIA / GC / STREP GROUP B ANTIGEN**

Chlamydia	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer
BD ProbeTec ET/Viper		23			23		22			23			23		
BD Max		3			3		3			3			3		
Cepheid		15			15		15			15			14	1	
Gen-Probe APTIMA		12			12		12			12			12		
Quidel QuickVue		5			5		4	1		5			5		
<b>Total Population</b>	<b>0</b>	<b>60</b>	<b>0</b>	<b>0</b>	<b>60</b>	<b>0</b>	<b>58</b>	<b>1</b>	<b>0</b>	<b>60</b>	<b>0</b>	<b>0</b>	<b>59</b>	<b>1</b>	<b>0</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>		<b>***</b>		<b>***</b>	<b>***</b>		<b>***</b>			<b>***</b>	<b>***</b>

Neisseria gonorrhoeae	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer
BD Max		3		3			3				3			3	
BD ProbeTec ET/Viper		23		23			22				23			23	
Cepheid		15		15			15				15			15	
Gen-Probe APTIMA	2	10		12			12				12			12	
<b>Total Population</b>	<b>2</b>	<b>53</b>	<b>0</b>	<b>55</b>	<b>0</b>	<b>0</b>	<b>54</b>	<b>0</b>	<b>0</b>	<b>0</b>	<b>55</b>	<b>0</b>	<b>55</b>	<b>0</b>	<b>0</b>
<b>Flagging</b>	<b>***</b>				<b>***</b>	<b>***</b>		<b>***</b>	<b>***</b>	<b>***</b>				<b>***</b>	<b>***</b>

Strep Group B	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Cepheid	3		3		3		3		3	
<b>Total Population</b>	<b>11</b>	<b>0</b>	<b>11</b>	<b>0</b>	<b>11</b>	<b>0</b>	<b>1</b>	<b>10</b>	<b>11</b>	<b>0</b>
<b>Flagging</b>		<b>***</b>		<b>***</b>		<b>***</b>	<b>***</b>			<b>***</b>

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.