



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2018**

**CRYPTOSPORIDIUM / GIARDIA IMMUNOASSAY**

***Cryptosporidium***

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Meridi ImmunoCard	7			7		7	7			7	7
Remel Xpect	3			3		3	3			3	3
<b>Total Population</b>	<b>15</b>	<b>0</b>	<b>1</b>	<b>14</b>	<b>0</b>	<b>15</b>	<b>15</b>	<b>0</b>	<b>1</b>	<b>14</b>	<b>15</b>
<b>Flagging</b>		***	***		***			***	***		

***Giardia lamblia***

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Meridi ImmunoCard	8		8			8	8		8		8
Remel Xpect	4		4			4	4		4		4
<b>Total Population</b>	<b>18</b>	<b>0</b>	<b>18</b>	<b>0</b>	<b>0</b>	<b>18</b>	<b>18</b>	<b>0</b>	<b>18</b>	<b>0</b>	<b>18</b>
<b>Flagging</b>		***		***	***			***		***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.