



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. w Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2018**

**RUBELLA**

Rubella IgG Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere Rubella-Plus		7	7		7		7		7		7
ASI Rubella Test		18	18		18		18		18		18
Beckman Access IgG		11	11		11		11		11		11
Biokit Rbgen/Fisher SureVue		5	5		5		5		5		5
Bio-Rad Bioplex 2000, 2200		4	4		4		4		4		4
DiaSorin Rubella IgG ELISA		3	3		3		3		3		3
Fisher Sure-Vue		7	7		7		6	1	7		7
Roche Elecsys		6	6		6		6		6		6
Siemens Immulite 2000, 2500		6	6		6		6		6		6
Siemens Immuno 1/Advia IgG		3	3		2		3		3		3
<b>Total Population</b>	<b>1</b>	<b>87</b>	<b>88</b>	<b>0</b>	<b>87</b>	<b>0</b>	<b>87</b>	<b>1</b>	<b>1</b>	<b>87</b>	<b>88</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>		<b>***</b>		<b>***</b>	<b>***</b>		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.