



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. w Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2018

STREP GROUP A ANTIGEN SCREEN

Step A Screen Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere Binax Now	4		4			4		4		4	4
Alere BioSta Acceava wvd	21		21			21		21		21	21
Alere i	6		6			6		6		6	6
BD Veritor	5		5			5		5		5	5
BTNX Rapid Response	3		3			3		3		3	3
Cardinal Health SP	8		8			8		8		8	8
Consult Dipstick, waived	14		14			14		14	2	12	14
Fisher Sure-Vue, waived	10		10			10		10		10	10
Henry Schein OneStep	3		3			3		3		3	3
Henry Schein OneStep+, Wv	4		4			4		4		4	4
McKesson, waived	6		6			6		6		6	6
Meridian Diag ImmunoCard	3		3			3		3		3	3
Quidel QuickVue +	18		18			18		18		18	18
Quidel QuickVue Dipstick	15		15			15	1	14		15	15
Quidel QuickVue In-Line	42		42			42		42	1	41	42
Quidel Sofia	9		9			9		9		9	9
Sekisui OSOM	8		8			8		8		8	8
Sekisui OSOM Ultra	12		12			12		12		12	12
Sekisui OSOM, waived	16		16			16		16		16	16
Total Population	223	0	223	0	0	223	1	222	3	220	223
Flagging		***		***	***		***		***		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.