



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

ToRCH

FIRST QUADRIMESTER 2018

Cytomegalovirus	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
Total Population	4	1	0	5	1	4
Flagging		***	***		***	

Herpes Type II or I/II, IgG	Specimen 1			Specimen 2			Specimen 3		
	Neg	Pos Type II	Pos for I or II	Neg	Pos Type II	Pos for I or II	Neg	Pos Type II	Pos for I or II
DiaSorin LIAISON	2	1			3		3		
Focus Diagnostics	1	2			3		3		
Total Population	10	3	0	0	13	0	13	0	0
Flagging				***				***	***

*Due to a lack of participant consensus, Specimen 1 was not evaluated. The intended result was negative.

Herpes Type I, IgG or non-spec	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
DiaSorin LIAISON	3			3	2	
Focus Diagnostics	3			3	2	1
Total Population	13	0	0	12	6	5
Flagging		***	***			

*Due to a lack of participant consensus, Specimen 3 was not evaluated this event. The intended result was Positive.

Herpes Type II or I/II, IgM	Specimen 1			Specimen 2			Specimen 3		
	Neg	Pos	Pos I/II	Neg	Pos	Pos I/II	Neg	Pos	Pos I/II
Total Population	6	0	1	5	1	0	5	0	0
Flagging		***	***		***	***		***	***

Herpes Type I, IgM	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
Total Population	3	0	3	0	3	0
Flagging		***		***		***

Rubella IgM	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
DiaSorin LIAISON	4			4	4	
Total Population	12	0	1	11	8	4
Flagging		***	***			

*Due to a lack of participant consensus, Specimen 3 was not graded. The intended result was Negative.

Toxoplasma	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
Total Population	6	0	6	0	0	6
Flagging		***		***	***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.