



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2018**

**URINE COLONY COUNT**

<b>Colony Count</b>	<b>Specimen 1</b>		<b>Specimen 2</b>	
	<b>No.</b>	<b>Flagging</b>	<b>No.</b>	<b>Flagging</b>
No Growth	0	***	33	
Growth of < 10,000 CFUs	14		11	
10,000 - 100,000 CFUs	29		0	***
>100,000 CFUs	1		0	***
<b>Total Population</b>	<b>44</b>			

\*Specimen 2 contained only normal skin flora. Therefore Specimen 2 was not evaluated.

<b>Identification</b>	<b>No.</b>	<b>Flagging</b>	<b>No.</b>	<b>Flagging</b>
<i>Staphylococcus epidermidis</i>	0	***	0	***
<i>Normal skin flora</i>	0	***	0	***
<i>Would refer/Test not performed</i>	11		8	
<i>Enteric, NOS</i>	0		0	***
<i>Corynebacterium, sp.</i>	0	***	1	
<i>Enterococcus sp.</i>	0	***	0	***
<i>Escherichia coli</i>	16		0	***
<i>Enterobacter cloacae</i>	0	***	0	***
<i>Klebsiella sp.</i>	0	***	0	***
<i>Proteus sp.</i>	0	***	0	***
<i>Pseudomonas sp.</i>	0	***	0	***
<i>Pseudomonas aeruginosa</i>	0	***	0	***
<i>Staphylococcus sp.</i>	0	***	1	***
<i>Staphylococcus saprophyticus</i>	0	***	0	***
No aerobic growth	0	***	7	
<b>Total Population</b>	<b>27</b>		<b>17</b>	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.