



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2018

VAGINOSIS SCREEN

Candida albicans

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
BD Affirm VPIII	18			18	17	1	1	17	18		18
Total Population	19	0	0	19	18	1	1	18	19	0	19
Flagging		***	***			***	***			***	

Gardnerella vaginalis

Method	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Labs
	BD Affirm VPIII	18			18	18		18		18	
Total Population	19	0	0	19	19	0	19	0	19	0	19
Flagging		***	***			***		***		***	

Trichomonas vaginalis

Method	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Labs
	BD Affirm VPIII		18		18	17	1	1	17	18	
Total Population	0	21	0	21	19	2	1	20	21	0	21
Flagging	***		***			***	***			***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.