



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. w Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

VIRAL ANTIGEN SCREEN

FIRST QUADRIMESTER 2018

Influenza Type A	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B
Alere		4		4			4			4			4		
Alere i		16		13	3		14	2		14	2		15	1	
BD Veritor		26		26			26			26			26		
Binax NOW Flu A&B		14		13	1		14			14			14		
Meridian Tru-Flu		3		3			3			3			3		
OraSure		16		16			16			16			16		
Quidel Quick Vue Infl A+B		7		7			7			7			7		
Quidel Sofia Infl A+B		22		22			22			22			22		
Remel Xpect Flu A&B	1	8		9			9			9			9		
Total Population	1	129	0	126	4	0	127	3	0	128	2	0	129	1	0
Flagging	***				***			***	***		***	***		***	***

Influenza Type B	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Alere i	15	1		16	12	4	14	2	14	2
BD Veritor	26			26	25	1	26		26	
Binax NOW Flu A&B	14			14	13	1	14		14	
OraSure	16			16	16		16		16	
Quidel Sofia Infl A+B	22			22	21	1	22		22	
Total Population	129	1	1	129	123	7	128	2	128	2
Flagging		***	***			***		***		***

Adenovirus	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Total Population	3	0	2	1	3	0	0	3	3	0
Flagging		***			***	***				***

*Due to a lack of participant consensus, Specimen 2 was not evaluated this event. The intended result was negative.

Respiratory Syncytial

Virus	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Alere Binax Now		26	26		26		26	26		
Alere i		7	7		7		7	7		
BD Veritor		3	3		3		3	3		
Binax NOW RSV		7	7		7		7	7		
Meridian Tru RSV		3	3		3		3	3		
Quidel QuickVue		9	9		9		9	9		
Quidel Sofia		8	8		8		8	8		
Total Population	1	73	74	0	74	0	0	74	74	0
Flagging	***			***		***	***			***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.