



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2018**

**VIRAL ANTIGEN SCREEN, Waived**

Influenza Type A	Specimen 1			Specimen 2		
	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B
Alere		18	1	17	1	
Alere i		18		17		
BD Veritor		55		47	1	
Binax NOW Flu A&B, wv		20	1	18	1	1
Henry Schein One Step +		4		4		
OraSure		5		5		
Quidel Quick Vue Infl A+B		11		9	1	
Quidel Sofia		17		17		
Sekisui OSOM A&B		6		5	1	
<b>Total Population</b>	<b>0</b>	<b>164</b>	<b>2</b>	<b>149</b>	<b>5</b>	<b>1</b>
<b>Flagging</b>	<b>***</b>				<b>***</b>	

Influenza Type B	Specimen 1		Specimen 2	
	Neg	Pos	Neg	Pos
Alere	17	2	1	19
Alere I	17			18
BD Veritor	47	2		54
Binax NOW Flu A&B, wv	19			20
Henry Schein One Step +	4			4
OraSure	5			5
Quidel Quick Vue Infl A+B	10			11
Quidel Sofia	17			17
Sekisui OSOM A&B	5	1		6
<b>Total Population</b>	<b>150</b>	<b>5</b>	<b>1</b>	<b>163</b>
<b>Flagging</b>		<b>***</b>	<b>***</b>	

Respiratory Syncytial Virus	Specimen 1		Specimen 2	
	Neg	Pos	Neg	Pos
Alere Binax Now		9		10
Alere i		4		4
BD Veritor		7		7
Quidel QuickVue		8		7
Quidel Sofia		3		3
<b>Total Population</b>	<b>0</b>	<b>33</b>	<b>33</b>	<b>1</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.