



Hematology With Diff G

Name	Line No.	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	
Leukocytes - Module G																	
Initial Grouping by Reagent																	
Coulter Ac*T 5	1	23.6 - 32.0	P 27.81	1.13	6.5 - 8.8	P 7.63	0.27	2.2 - 3.0	P 2.62	0.1	4.6 - 6.2	P 5.38	0.16	14.6 - 19.8	P 17.19	0.64	18
Horiba ABX Pentra 60 C+	2	23.8 - 32.2	P 27.98	0.62	6.4 - 8.7	P 7.57	0.21	2.2 - 3.0	P 2.61	0.07	4.5 - 6.1	P 5.34	0.11	14.5 - 19.7	P 17.11	0.25	17
Horiba ABX Pentra 80	3	23.6 - 32.0	P 27.79	0.97	6.6 - 8.9	P 7.71	0.16	2.3 - 3.1	P 2.69	0.09	4.7 - 6.3	P 5.49	0.11	14.6 - 19.8	P 17.2	0.33	16
Initial Grouping by Sensitivity or Principle																	
Pentra 60 / Ac*T 5	4	23.7 - 32.1	P 27.92	0.88	6.5 - 8.7	P 7.6	0.24	2.2 - 3.0	P 2.61	0.09	4.6 - 6.2	P 5.36	0.14	14.6 - 19.7	P 17.16	0.48	38
Other Instruments	5	23.6 - 32.0	P 27.79	0.97	6.6 - 8.9	P 7.71	0.16	2.3 - 3.1	P 2.69	0.09	4.7 - 6.3	P 5.49	0.11	14.6 - 19.8	P 17.2	0.33	16
Total Population																	
Whole Population	6	23.7 - 32.1	P 27.88	0.91	6.5 - 8.8	P 7.63	0.23	2.2 - 3.0	P 2.63	0.09	4.6 - 6.2	P 5.39	0.15	14.6 - 19.7	P 17.17	0.44	54
Erythrocytes - Module G																	
Initial Grouping by Reagent																	
Coulter Ac*T 5	1	2.8 - 3.16	P 2.983	0.06	4.25 - 4.8	P 4.526	0.044	1.94 - 2.19	P 2.062	0.043	3.96 - 4.46	P 4.211	0.055	5.32 - 6.0	P 5.656	0.087	18
Horiba ABX Pentra 60 C+	2	2.78 - 3.14	P 2.961	0.054	4.2 - 4.73	P 4.465	0.099	1.93 - 2.17	P 2.05	0.043	3.89 - 4.38	P 4.134	0.093	5.24 - 5.9	P 5.57	0.1	16
Horiba ABX Pentra 80	3	2.75 - 3.1	P 2.929	0.048	4.19 - 4.73	P 4.46	0.072	1.88 - 2.12	P 2.003	0.047	3.89 - 4.38	P 4.133	0.055	5.27 - 5.94	P 5.606	0.084	16
Initial Grouping by Sensitivity or Principle																	
Pentra 60 / Ac*T 5	4	2.79 - 3.15	P 2.97	0.058	4.22 - 4.76	P 4.491	0.081	1.93 - 2.18	P 2.055	0.045	3.92 - 4.42	P 4.173	0.087	5.27 - 5.94	P 5.606	0.106	37
Other Instruments	5	2.75 - 3.1	P 2.929	0.048	4.19 - 4.73	P 4.46	0.072	1.88 - 2.12	P 2.003	0.047	3.89 - 4.38	P 4.133	0.055	5.27 - 5.94	P 5.606	0.084	16
Total Population																	
Whole Population	6	2.78 - 3.13	P 2.957	0.058	4.21 - 4.75	P 4.481	0.079	1.92 - 2.16	P 2.039	0.051	3.91 - 4.41	P 4.159	0.081	5.27 - 5.94	P 5.603	0.101	54
Hemoglobin - Module G																	
Initial Grouping by Reagent																	
Coulter Ac*T 5	1	7.0 - 8.1	P 7.57	0.08	12.2 - 14.1	P 13.17	0.11	6.4 - 7.3	P 6.86	0.1	11.2 - 12.9	P 12.07	0.12	15.1 - 17.4	P 16.24	0.12	18
Horiba ABX Pentra 60 C+	2	6.9 - 8.0	P 7.47	0.12	12.3 - 14.1	P 13.18	0.17	6.3 - 7.2	P 6.73	0.12	11.2 - 12.9	P 12.04	0.19	15.2 - 17.5	P 16.34	0.22	17
Horiba ABX Pentra 80	3	6.9 - 7.9	P 7.41	0.09	12.2 - 14.0	P 13.13	0.14	6.2 - 7.1	P 6.65	0.1	11.1 - 12.8	P 11.96	0.11	15.2 - 17.5	P 16.31	0.16	15
Initial Grouping by Sensitivity or Principle																	
Pentra 60 / Ac*T 5	4	7.0 - 8.0	P 7.51	0.11	12.2 - 14.1	P 13.16	0.15	6.3 - 7.3	P 6.78	0.13	11.2 - 12.9	P 12.04	0.17	15.1 - 17.4	P 16.29	0.2	39
Other Instruments	5	6.9 - 7.9	P 7.41	0.09	12.2 - 14.0	P 13.13	0.14	6.2 - 7.1	P 6.65	0.1	11.1 - 12.8	P 11.96	0.11	15.2 - 17.5	P 16.31	0.16	15
Total Population																	
Whole Population	6	7.0 - 8.0	P 7.48	0.12	12.2 - 14.1	P 13.15	0.15	6.3 - 7.2	P 6.74	0.14	11.2 - 12.9	P 12.02	0.16	15.1 - 17.4	P 16.29	0.19	54
Hematocrit - Module G																	
Initial Grouping by Reagent																	
Coulter Ac*T 5	1	20.5 - 23.1	P 21.81	0.48	35.2 - 39.7	P 37.48	0.57	17.8 - 20.1	P 18.93	0.35	32.2 - 36.3	P 34.26	0.48	43.7 - 49.2	P 46.45	0.8	18
Horiba ABX Pentra 60 C+	2	20.3 - 22.9	P 21.62	0.33	35.4 - 39.9	P 37.61	0.75	17.7 - 20.0	P 18.85	0.34	31.9 - 36.0	P 33.95	0.76	43.7 - 49.3	P 46.47	0.93	17
Horiba ABX Pentra 80	3	20.7 - 23.4	P 22.06	0.37	35.3 - 39.8	P 37.59	0.56	18.0 - 20.3	P 19.16	0.44	32.1 - 36.2	P 34.18	0.44	43.4 - 48.9	P 46.15	0.64	16
Initial Grouping by Sensitivity or Principle																	
Pentra 60 / Ac*T 5	4	20.4 - 23.0	P 21.71	0.44	35.3 - 39.8	P 37.51	0.69	17.7 - 20.0	P 18.88	0.4	32.1 - 36.1	P 34.1	0.71	43.6 - 49.2	P 46.41	0.93	38
Other Instruments	5	20.7 - 23.4	P 22.06	0.37	35.3 - 39.8	P 37.59	0.56	18.0 - 20.3	P 19.16	0.44	32.1 - 36.2	P 34.18	0.44	43.4 - 48.9	P 46.15	0.64	16
Total Population																	
Whole Population	6	20.5 - 23.1	P 21.81	0.45	35.3 - 39.8	P 37.53	0.66	17.8 - 20.1	P 18.96	0.43	32.1 - 36.2	P 34.12	0.64	43.6 - 49.1	P 46.33	0.86	54
Platelets - Module G																	
Initial Grouping by Reagent																	
Coulter Ac*T 5	1	284 - 474	P 379.2	14.5	184 - 306	P 244.8	8.9	55 - 91	P 73.1	5.1	184 - 306	P 245.2	9.4	224 - 373	P 298.1	9.2	18
Horiba ABX Pentra 60 C+	2	284 - 474	P 379.2	12.6	182 - 303	P 242.4	12.0	54 - 89	P 71.4	3.4	183 - 306	P 244.5	8.8	219 - 365	P 291.8	10.1	16

Horiba ABX Pentra 80	3	286 - 476	P 381.1	11.2	185 - 309	P 247.2	10.1	54 - 90	P 72.4	3.9	188 - 313	P 250.0	10.3	233 - 389	P 310.8	9.5	16
Initial Grouping bySensitivityor Principle																	
Pentra 60 / Ac*T 5	4	285 - 475	P 379.7	13.3	183 - 306	P 244.4	10.9	55 - 91	P 72.8	4.9	184 - 307	P 245.5	9.2	222 - 371	P 296.6	11.2	37
Other Instruments	5	286 - 476	P 381.1	11.2	185 - 309	P 247.2	10.1	54 - 90	P 72.4	3.9	188 - 313	P 250.0	10.3	233 - 389	P 310.8	9.5	16
Total Population																	
Whole Population	6	285 - 475	P 380.2	12.6	184 - 306	P 245.2	10.6	54 - 91	P 72.6	4.6	185 - 308	P 246.8	9.7	225 - 376	P 300.6	12.7	54

Neutrophil % - Module G

Initial Grouping byReagent																	
Coulter Ac*T 5	1	56.0 - 69.1	S 62.53	2.18	48.0 - 63.6	S 55.82	2.6	51.8 - 71.5	S 61.64	3.27	50.0 - 71.2	S 60.6	3.54	38.4 - 56.4	S 47.36	3.0	18
Horiba ABX Pentra 60 C+	2	54.8 - 78.0	S 66.4	3.87	49.7 - 70.5	S 60.12	3.47	55.4 - 75.1	S 65.27	3.29	56.2 - 77.0	S 66.61	3.47	44.7 - 62.0	S 53.36	2.89	17
Horiba ABX Pentra 80	3	58.4 - 74.5	S 66.44	2.68	47.4 - 69.4	S 58.39	3.66	55.6 - 75.9	S 65.78	3.39	47.4 - 78.0	S 62.68	5.1	34.0 - 59.4	S 46.73	4.24	16
Initial Grouping bySensitivityor Principle																	
Pentra 60 / Ac*T 5	4	53.2 - 76.2	S 64.7	3.84	46.9 - 69.0	S 57.96	3.67	52.3 - 75.2	S 63.72	3.82	50.1 - 77.4	S 63.74	4.55	38.2 - 62.7	S 50.46	4.08	38
Other Instruments	5	58.4 - 74.5	S 66.44	2.68	47.4 - 69.4	S 58.39	3.66	55.6 - 75.9	S 65.78	3.39	47.4 - 78.0	S 62.68	5.1	34.0 - 59.4	S 46.73	4.24	16
Total Population																	
Whole Population	6	54.4 - 76.1	S 65.22	3.62	47.1 - 69.1	S 58.09	3.67	52.9 - 75.8	S 64.34	3.81	49.2 - 77.7	S 63.42	4.75	35.9 - 62.7	S 49.31	4.47	54

Lymphocyte % - Module G

Initial Grouping byReagent																	
Coulter Ac*T 5	1	13.9 - 26.9	S 20.39	2.17	21.5 - 37.0	S 29.22	2.58	10.0 - 30.5	S 20.29	3.41	15.1 - 37.0	S 26.03	3.66	28.4 - 50.3	S 39.36	3.64	18
Horiba ABX Pentra 60 C+	2	13.2 - 22.6	S 17.92	1.57	20.2 - 31.1	S 25.64	1.81	10.0 - 26.4	S 18.16	2.73	12.7 - 29.1	S 20.9	2.73	27.8 - 41.4	S 34.6	2.26	17
Horiba ABX Pentra 80	3	15.6 - 35.6	S 25.59	3.33	19.7 - 47.2	S 33.45	4.58	14.0 - 32.1	S 23.04	3.01	12.5 - 46.4	S 29.45	5.66	30.3 - 60.0	S 45.13	4.94	16
Initial Grouping bySensitivityor Principle																	
Pentra 60 / Ac*T 5	4	12.3 - 26.5	S 19.43	2.37	18.8 - 36.8	S 27.79	2.99	9.5 - 29.3	S 19.4	3.29	11.3 - 36.0	S 23.63	4.11	25.7 - 48.7	S 37.21	3.83	38
Other Instruments	5	15.6 - 35.6	S 25.59	3.33	19.7 - 47.2	S 33.45	4.58	14.0 - 32.1	S 23.04	3.01	12.5 - 46.4	S 29.45	5.66	30.3 - 60.0	S 45.13	4.94	16
Total Population																	
Whole Population	6	9.6 - 33.0	S 21.29	3.91	16.3 - 42.7	S 29.5	4.39	9.6 - 31.4	S 20.5	3.62	9.3 - 41.4	S 25.39	5.35	22.9 - 56.4	S 39.65	5.57	54

Monocyte % - Module G

Initial Grouping byReagent																	
Coulter Ac*T 5	1	0.1 - 3.3	S 1.69	0.53	0.4 - 3.7	S 2.08	0.56	0 - 1.6	S 0.62	0.34	0 - 2.3	S 0.87	0.47	0 - 3.4	S 1.68	0.57	18
Horiba ABX Pentra 60 C+	2	0 - 3.2	S 1.52	0.56	0 - 3.6	S 1.83	0.6	0 - 1.6	S 0.56	0.36	0 - 2.0	S 0.77	0.4	0 - 3.1	S 1.21	0.62	17
Horiba ABX Pentra 80	3	0 - 2.7	S 1.32	0.47	0 - 3.8	S 1.64	0.72	0 - 1.7	S 0.48	0.41	0 - 2.4	S 0.75	0.55	0 - 2.7	S 1.09	0.52	16
Initial Grouping bySensitivityor Principle																	
Pentra 60 / Ac*T 5	4	0 - 3.2	S 1.63	0.53	0.2 - 3.7	S 1.96	0.57	0 - 1.6	S 0.6	0.34	0 - 2.2	S 0.86	0.44	0 - 3.3	S 1.46	0.62	38
Other Instruments	5	0 - 2.7	S 1.32	0.47	0 - 3.8	S 1.64	0.72	0 - 1.7	S 0.48	0.41	0 - 2.4	S 0.75	0.55	0 - 2.7	S 1.09	0.52	16
Total Population																	
Whole Population	6	0 - 3.1	S 1.54	0.53	0 - 3.8	S 1.87	0.64	0 - 1.7	S 0.56	0.37	0 - 2.2	S 0.82	0.47	0 - 3.2	S 1.35	0.61	54

Eosinophil % - Module G

Initial Grouping byReagent																	
Coulter Ac*T 5	1	2.2 - 5.2	S 3.69	0.51	2.4 - 6.2	S 4.31	0.63	2.0 - 14.4	S 8.18	2.06	2.7 - 8.0	S 5.34	0.88	3.3 - 8.8	S 6.03	0.92	18
Horiba ABX Pentra 60 C+	2	2.5 - 5.5	S 3.98	0.51	2.7 - 6.8	S 4.78	0.69	1.4 - 13.9	S 7.67	2.07	1.9 - 9.4	S 5.63	1.25	3.5 - 9.0	S 6.26	0.91	16
Horiba ABX Pentra 80	3	2.6 - 5.3	S 3.94	0.44	1.9 - 6.9	S 4.43	0.83	2.3 - 14.2	S 8.23	1.98	1.7 - 8.1	S 4.9	1.06	1.9 - 8.9	S 5.42	1.17	16
Initial Grouping bySensitivityor Principle																	
Pentra 60 / Ac*T 5	4	2.2 - 5.4	S 3.81	0.53	2.5 - 6.6	S 4.55	0.68	1.9 - 13.9	S 7.9	2.01	2.3 - 8.6	S 5.45	1.04	3.1 - 9.0	S 6.03	0.98	37
Other Instruments	5	2.6 - 5.3	S 3.94	0.44	1.9 - 6.9	S 4.43	0.83	2.3 - 14.2	S 8.23	1.98	1.7 - 8.1	S 4.9	1.06	1.9 - 8.9	S 5.42	1.17	16
Total Population																	
Whole Population	6	2.3 - 5.4	S 3.85	0.51	2.3 - 6.7	S 4.52	0.73	2.0 - 14.0	S 8.0	2.01	2.0 - 8.5	S 5.28	1.08	2.6 - 9.1	S 5.84	1.08	53

Basophil % - Module G

Initial Grouping byReagent																	
Coulter Ac*T 5	1	9.1 - 14.3	S 11.7	0.85	6.8 - 10.5	S 8.65	0.63	7.8 - 10.6	S 9.2	0.45	6.2 - 8.2	C 7.16	0.32	3.7 - 7.5	S 5.58	0.62	18
Horiba ABX Pentra 60 C+	2	0 - 24.5	S 10.25	4.76	0 - 18.2	S 7.62	3.52	0 - 18.6	S 7.78	3.6	0 - 14.4	S 6.12	2.76	0 - 11.0	S 4.63	2.13	16
Horiba ABX Pentra 80	3	0 - 13.8	S 2.71	3.71	0 - 10.9	S 2.09	2.93	0 - 13.0	S 2.48	3.51	0 - 10.8	S 2.06	2.9	0 - 8.6	S 1.64	2.33	16
Initial Grouping bySensitivityor Principle																	
Pentra 60 / Ac*T 5	4	0 - 22.5	S 10.47	4.02	0 - 16.7	S 7.77	2.99	0 - 17.4	S 8.11	3.09	0 - 13.5	S 6.34	2.37	0 - 10.6	S 4.86	1.91	37
Other Instruments	5	0 - 13.8	S 2.71	3.71	0 - 10.9	S 2.09	2.93	0 - 13.0	S 2.48	3.51	0 - 10.8	S 2.06	2.9	0 - 8.6	S 1.64	2.33	16
Total Population																	

Whole Population	6	0 - 24.0	S 8.13	5.31	0 - 17.9	S 6.06	3.95	0 - 18.8	S 6.41	4.13	0 - 14.7	S 5.05	3.21	0 - 11.5	S 3.87	2.53	53
------------------	---	----------	--------	------	----------	--------	------	----------	--------	------	----------	--------	------	----------	--------	------	----