



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2019**

**ACID-FAST SMEARS**

Results	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Absent	Present	Absent	Present	Absent	Present	Absent	Present	Absent	Present
<b>Total Population</b>	8	0	0	8	8	0	0	8	8	0
<b>Flagging</b>		***	***			***	***			***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.