



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**
205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2019

CHLAMYDIA / GC / STREP GROUP B ANTIGEN

Chlamydia	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	
BD Max	3			3			3				3			3		
BD ProbeTec ET/Viper	20			19	1		20				20			20		
Cepheid	19			19			19				19			19		
Gen-Probe APTIMA	11			11			11				11			11		
Quidel QuickVue	4	1		5			4	1		2	3		4	1		
Total Population	58	1	0	58	1	0	58	1	0	2	57	0	4	55	0	
Flagging		***			***	***		***		***			***			

Neisseria gonorrhoeae	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	
BD Max		3		3				3		3				3		
BD ProbeTec ET/Viper		20		19	1			20		20				20		
Cepheid		19		19				19		19				19		
Gen-Probe APTIMA	3	8		11			1	10		11			11			
Total Population	3	51	0	53	1	0	1	53	0	54	0	0	54	0	0	
Flagging	***				***	***	***				***			***		

Strep Group B	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
BD Max	3		3		3		3		3	
Cepheid	5		5		5		5		5	
Total Population	11	0	11	0	11	0	1	10	1	10
Flagging		***		***		***	***		***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.