



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2019**

**CRYPTOSPORIDIUM / GIARDIA IMMUNOASSAY**

***Cryptosporidium***

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Cardinal Health		3	3		3		2	1		3	3
Meridi ImmunoCard		6	6		6		6			6	6
Remel Xpect		3	3		3		3			3	3
<b>Total Population</b>	<b>0</b>	<b>14</b>	<b>14</b>	<b>0</b>	<b>14</b>	<b>0</b>	<b>13</b>	<b>1</b>	<b>0</b>	<b>14</b>	<b>14</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>		<b>***</b>		<b>***</b>	<b>***</b>		

***Giardia lamblia***

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Cardinal Health	3		3		3			3	3		3
Meridi ImmunoCard	7		7		7			7	7		7
Remel Xpect	3		3		3			3	3		3
<b>Total Population</b>	<b>15</b>	<b>0</b>	<b>15</b>	<b>0</b>	<b>15</b>	<b>0</b>	<b>0</b>	<b>15</b>	<b>15</b>	<b>0</b>	<b>15</b>
<b>Flagging</b>		<b>***</b>		<b>***</b>		<b>***</b>	<b>***</b>			<b>***</b>	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.