



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2019

C-Reactive Protein, qualitative

C-REACTIVE PROTEIN Method	Specimen 1		Specimen 2		No. of Labs
	NEG	POS	NEG	POS	
ASI	7	13		20	20
Immunostics immuno/lex-crp		5	1	4	5
Stanbio #1130		3		3	3
Teco Diag latex slide test		4		4	4
Total Population	8	31	1	38	39
Flagging	***		***		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.