



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2019

HELICOBACTER PYLORI

<i>H. pylori</i>	Specimen 1		Specimen 2		No. of Labs
	Neg	Pos	Neg	Pos	
Alere Clearview	12			12	12
Alere ELISA	3			3	3
Alfa Scientific InstantView	20			20	20
BTNX Rapid Response, wv	6			6	6
Cardinal Health SP	5			5	5
Consult Diagnostics	6			6	6
Fisher Sure-Vue IgG	7			7	7
Henry Schein One Step	8			8	8
Micro Detect IgG	3			3	3
Quidel QuickVue gII	18	1	1	18	19
Total Population	110	1	2	109	111
Flagging		***	***		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.