



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2019

Methicillin-resistant Staphylococcus aureus (MRSA)

Method	Specimen 1		Specimen 2	
	Neg	Pos	Neg	Pos
Total Population	8	0	0	8
Flagging		***	***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees.

Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.