



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**
205 West Levee St. w Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2019

RUBELLA

Rubella IgG Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere Rubella-Plus		6	6		6		5	1		6	6
ASI Rubella Test		16	16		16		15	1		16	16
Beckman Access IgG		11	11		11		11		1	10	11
Biokit Rbgen/Fisher SureVue		4	4		4		4			4	4
Bio-Rad Bioplex 2000, 2200		4	4		4		4			4	4
Fisher Sure-Vue		6	6		6		6			6	6
Roche Elecsys		5	5		5		5			5	5
Siemens Immulite 2000, 2500		6	6		6		6			6	6
Siemens Immuno 1/Advia IgG		4	4		4		4			4	4
Total Population	0	81	80	1	80	1	78	3	1	80	81
Flagging	***			***		***		***	***		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.