



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. w Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2019

STREP GROUP A ANTIGEN SCREEN

Strep A Screen Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere BioSta Acceava wvd		7	7		7			7		7	7
Alere i		14	14		14			14		14	14
BD Veritor		4	4		4			4		4	4
BTNX Rapid Response		3	3		3			3		3	3
Cardinal Health SP		4	4		4			4		4	4
Cepheid GeneXpert		7	7		7			7		7	7
Consult Dipstick, waived	1	6	7		7			7		7	7
Fisher Sure-Vue, waived	1	9	10		10			10		10	10
Henry Schein OneStep+, Wv		3	3		3			3		3	3
McKesson, waived		9	8		8			9		9	9
Meridian Diag ImmunoCard		4	4		4			4		4	4
Quidel QuickVue +		7	7		7			7		7	7
Quidel QuickVue Dipstick		13	13		13			13		13	13
Quidel QuickVue In-Line	1	32	32	1	32	1		33	1	32	33
Quidel Sofia		8	8		8			8		8	8
Sekisui OSOM		7	7		7			7		7	7
Sekisui OSOM Ultra		15	15		15			15		15	15
Sekisui OSOM, waived		23	23		23			23		23	23
Sekisui Rapid Strep A		7	7		7			7		7	7
Total Population	3	194	195	1	195	1	0	197	1	196	197
Flagging	***			***		***	***		***		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.